



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 07001F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	230926.1 DT - 26 / 09 / 2023			ULR No.	CC2646230000 07001F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	26.09.2023		
Address	DH Civil Hospital, Jalandhar-Hoshiarpur Road, Hoshiarpur (PB)			Recommended Date for Next Callb.	25.09.2024		
				Certificate Issue Dt.	29.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI. No	Party ID / Location
1	Centrifuge M/c.	0 to 6000 RPM	10 RPM	Remi -----	KDDHPBCPCF07 Bio Chemistry

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	(50 ± 20)%RH	(25 ± 10) °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

Calibration Result

SI	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in RPM)
	UUC (in RPM)	(in RPM)	(in RPM)	
1	100	100.1	-0.1	5.88
2	500	500.2	-0.2	5.88
3	1000	1000.2	-0.2	5.88
4	3000	3000.4	-0.4	6.15
5	6000	6000.6	-0.6	6.17

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Anand Chaurasia
(CEO)