



# A A CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



|                      |                                                               |                     |       |                                  |                     |          |        |
|----------------------|---------------------------------------------------------------|---------------------|-------|----------------------------------|---------------------|----------|--------|
| Certificate No       | AACPL / 07008F                                                | Field               | Mech. | Group                            | Volume              | Page No. | 1 of 1 |
| SRF No & Date        | 230926.2                                                      | DT - 26 / 09 / 2023 |       | ULR No.                          | CC2646230000 07008F |          |        |
| Calibration Method : | Comparison Method                                             |                     |       | Accuracy                         |                     |          |        |
| Company Name         | M/s Krsnaa Diagnostics Ltd.                                   |                     |       | Calibration Date                 | 26.09.2023          |          |        |
| Address              | DH Civil Hospital, Jalandhar-Hoshiarpur Road, Hoshiarpur (PB) |                     |       | Recommended Date for Next Callb. | 25.09.2024          |          |        |
|                      |                                                               |                     |       | Certificate Issue Dt.            | 29.09.2023          |          |        |
|                      |                                                               |                     |       | UUC Condition                    | Good                |          |        |

### Calibration Instrument Detail

| Sl | Instrument Name | Range        | Least Count | Make / SI.No.    | Party ID / Location           |
|----|-----------------|--------------|-------------|------------------|-------------------------------|
| 1  | Micro Pipette   | 10 to 100 µl | 1 µl        | I-Pitte<br>----- | KDDHPBCCMP11<br>Bio Chemistry |

### Standard Equipments Used ( Traceable to National Standard )

| Sl | Instrument Name  | Make/SI No      | Calibrated By  | Cal Certificate No | Cal. Valid Upto |
|----|------------------|-----------------|----------------|--------------------|-----------------|
| 1  | Weighing Balance | Radwag / 668683 | Bluebox Techno | BBT/452/JAN/23     | 05/03/2024      |

| Reference Standard | Calibration Purpose | Humidity     | Temperature | Calibration Performed At |
|--------------------|---------------------|--------------|-------------|--------------------------|
| ISO - 8655 - 6     | Volume Measurement  | (50 ± 10)%RH | (23 ± 2)°C  | Laboratory               |

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

### Calibration Result

| S. No | Nominal Value on UUC (in µl) | Average Volume at 27 °C (in µl) | Deviation (in µl) | Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is ± (in µl) |
|-------|------------------------------|---------------------------------|-------------------|-------------------------------------------------------------------------------------|
| 1     | 10                           | 10.00259                        | -0.00259          | 0.25                                                                                |
| 2     | 50                           | 50.00563                        | -0.00563          | 0.25                                                                                |
| 3     | 100                          | 100.00869                       | -0.00869          | 0.25                                                                                |

### Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

*Asbich*  
Calibrated by



Approved by  
*Arvind Chaurasia*  
Arvind Chaurasia  
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012  
Ph. : 0120-4135529, Mob. : 9212172564, 9212172597  
E-mail : aacplinfo2005@gmail.com, accplinfo@rediffmail.com