



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 07009F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	230926.2	DT - 26 / 09 / 2023		ULR No.	CC2646230000 07009F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	26.09.2023		
Address	DH Civil Hospital, Jalandhar-Hoshiarpur Road, Hoshiarpur (PB)			Recommended Date for Next Callib	25.09.2024		
				Certificate Issue Dt.	29.09.2023		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / Sl.No.	Party ID / Location
1	Micro Pipette	5 to 50 μ l	1 μ l	I-Pitte CU02274015	KDDHPBCCMP13 Bio Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 \pm 10)%RH	(23 \pm 2) $^{\circ}$ C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in μ l)	Average Volume at 27 $^{\circ}$ C (in μ l)	Deviation (in μ l)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in μ l)
1	10	9.99810	0.00190	0.25
2	30	29.98614	0.01386	0.25
3	50	49.98120	0.01880	0.25

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Agulsh
Calibrated by



Approved by
Agulsh
Anand Chaurasia
(CEO)