



# The Accurate Calibration & Medical Solution

## CALIBRATION CERTIFICATE

### CERTIFICATE OF PERFORMANCE FOR MICROSCOPE

ISSUED TO: M/S KRSNAA DIAGNOSTICS LTD.  
ADDRESS: DH CIVIL HOSPITAL, JALANDHAR - HOSHIARPUR RAOD, HOSHIARPUR (PB) - 146001  
CERTIFICATE NO.: TAC/09-23/KDL/MS-1001  
LOCATION/DEPT.: CLINICAL PATHOLOGY  
ID NO./BAR CODE: KDDHPBCPMS05  
MAKE: GLOMAX

DATE: 26.09.2023  
MODEL NO.: LX-300  
SERIAL NO.: NA

### CALIBRATION STANDARD USED

Sr.No.	Master Equipment	Manufacturer	Model No.	Serial No.	Cal. Date	Cal Due Date
1	Electrical Safety Analyzer	DATREND Systems Inc.	vPed - Mini	VPM18100028	15.10.2022	15.10.2023

### ENVIRONMENTAL CONDITIONS

Temperature :  $22 \pm 2^{\circ}\text{C}$   
Humidity :  $50 \pm 10\% \text{RH}$

### VISUAL TEST

Sr.No.	Tests	Remarks
1	Power Cords, Cable Checking	OK
2	Mains Socket Checking	OK
3	Equipment Type (B, BF, CF)	CF
4	Equipment Class (I, II, III)	I

### FUNCTIONAL TEST

Sr.No.	Parameters	Measured Value	Remarks
1	Voltage Between Live and Neutral (VLN)	231.2 V (220-240 V)	OK
2	Voltage Between Live and Earth (VLE)	230.2 V (220-240V)	OK
3	Voltage Between Neutral and Earth (VNE)	0.4 V (0-5V)	OK
4	Load Current (Ia)	0.4 A	OK
5	Enclosure Current	0.6 $\mu\text{A}$ (0-10 $\mu\text{A}$ )	OK
6	Leaking Current (IL)	142 $\mu\text{A}$ (0-300 $\mu\text{A}$ )	OK
7	Time for Testing (t)	15 (min)	OK

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed YES  NO

Calibration Date: 26. 09. 2023

Calibration Due Date: 25. 09. 2024

### Remarks:

- Certificates shall not be reproduced except in full, without the written approval of The Accurate Calibration & Medical Solution.
- The Calibration result reported in this certificate are valid at the time of calibration and the stated condition of measurement.

Calibrated by



Approved by

Plot No. 72/G1, Ganesh Vihar Sirsi Road, Jaipur (Raj.) -302012

Mob.: 9672294758

E-mail : theaccuratecalibration@gmail.com