



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2648

Certificate No	AACPL / 07002F	Field	Thermal	Group	Temp.	Page No.	1 of 1
SRF No & Date	230926.1	DT - 26 / 09 / 2023		ULR No.		CC2646230000	07002F
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date		26.09.2023	
Address	DH Civil Hospital, Jalandhar-Hoshiarpur Road, Hoshiarpur (PB)			Recommended Date for Next Calib.		25.09.2024	
				Certificate Issue Dt.		29.09.2023	
				UUC Condition		Good	

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make/SI.No.	Party ID / Location
1	Hot Air Oven	0 to 300 °C	0.1 °C	Arba Healthcare	KDDHPBSEIN12 Serology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Digital Multimeter	Gwinstek EL 160577	AACPL	AACPL/04256F	11/07/2024
2	RTD (4 Wire) Pt-100	Tempens/ 115	AACPL	AACPL/19820F	28/09/2023

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
DKD-R5-7	Temp. Measurement	(50 ± 20)%RH	(25 ± 10)°C	Site

Calibration Procedure

Comparison With Standard Sensors as per Calibration Procedure Manual AACPL / CP / 02, Section No : CPTH - 09

Calibration Points :- 10 , 150 & 300 °C

Calibration Results				
SI No.	Master Value	UUC Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor k =2 is ± (°C)
	In °C	In °C	In °C	
1	10.05	10.2	0.15	0.32
2	150.09	150.3	0.21	0.32
3	300.13	299.7	-0.43	0.84

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration
- Temperature Scale: International Temperature Scale Of 1990 (ITS-90)

----- End of Certificate -----

Calibrated by



Approved by

Anand Chaurasia
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012

Ph. : 0120-4135529, Mob. : 9212172564, 9212172597

E-mail : aacplinfo2005@gmail.com, acplinfo@rediffmail.com