

Certification of Calibration

Instrument Name	: Automated 3 Part Cell Counter
Make	: Nihon Kohden
Model	: MEK-1305
Serial no	: 17
Calibration date	: 18-Oct-23
Installation place	: Velys diagnostics & pathology center private limited

This is to certify that the above mentioned instrument has been successfully calibrated on **17th Oct 2023** with MEK-CAL lot no-**PLUS-23X** bearing expiry till **05th Nov 2023**. During the calibration of the analyser, all CBC parameters were calibrated.

Run Quality Controls **MEK-3D (Low, Normal, and High) Lot no-B238** bearing expiry of **05th Nov 2023**, Result found within specified range.

Based on the manufacture recommended calibration interval the next due date of calibration is on **17th Oct 2024**. Validity \pm 30days. Or depends on QC performance/subject to replacement or change in hardware.

For Nihon Kohden India Pvt.
Ltd.

Name : ANKIT DAVE
Designation : CS

