



Metropolis Healthcare Ltd., EQAS
VIOMA LIFESCIENCES PVT LTD(FORMERLY HWELL24 PLUS
HEALTHCARE PVT LTD) (1013)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit
X Glucose (urine)	Urometer 120	1+	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____


Reviewed by: _____

Dated: _____

Instrument : Urometer 120

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6.5	6,6.5,7
@ Specific Gravity	-	1.025	1.015,1.020,1.025
X Glucose (urine)	-	1+	+2,+3,+4
@ Protein (Urine)	-	2+	+1,+2,+3
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	3+	+1,+2,+3
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable	Total Parameters	9
@ : Acceptable	Not Evaluated Parameters	0
X : Unacceptable	Evaluated Parameters	9
# : Not Evaluated	Outlier Parameters (X)	1
Ⓢ : Delayed Result Entry	EQAS Score Clinical Path - Urine Routine Chemistry	88.89 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070

CORRECTIVE & PREVENTIVE ACTIONS

ROOT CAUSE ANALYSIS, CORRECTIVE AND PREVENTIVE ACTION (CAPA)

Date: 22/12/23

Non Compliance Reference

No. 2 dated 19/12/23
 Dept: C.P. (urine glucose)

Incident	Used Expired urine analyzing strips. The technician who did was fresher as there were only few strips in the packet, he was not able to check expiry date, making the test result go wrong.
Root cause analysis/Investigation	strip was expired which was used for testing, it made interpretation of glucose wrong. Hence the result was compromised. Done By: Akshay Date & Sign: <u>Ajmal</u> 22/12/23
Tool Used/ Risk Assessment	Brainstorming
Immediate Correction	Discarded Expire strips.
Corrective Action	When we found that strips were expired, we replaced them with new ones and told the same to the department technician. Done By: Akshay Date & Sign: <u>Ajmal</u> 22/12/23
Preventive Action	Should check exping of all kits / tools used for the tests before starting the test.
Verification of CAPA Done	Done By: Dr Kupriya Date & Sign: <u>m</u> 23/12/23

Akshay Ajmal
 Prepared By

Ajmal
 Immediate Incharge

m
 HOD



Metropolis Healthcare Ltd., EQAS
 VIOMA LIFESCIENCES PVT LTD(FORMERLY HWELL24 PLUS
 HEALTHCARE PVT LTD) (1013)
 Quality And Analyte Summary Report
 Result Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Instrument : Urometer 120

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6.5	6,6.5,7
@ Specific Gravity	-	1.025	1.015,1.020,1.025
X Glucose (urine)	-	1+	+2,+3,+4
@ Protein (Urine)	-	2+	+1,+2,+3
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	3+	+1,+2,+3
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable	Total Parameters	9
⊖ : Acceptable	Not Evaluated Parameters	0
X : Unacceptable	Evaluated Parameters	9
# : Not Evaluated	Outlier Parameters (X)	1
⌚ : Delayed Result Entry	EQAS Score Clinical Path - Urine Routine Chemistry	88.89 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W),
 Mumbai - 400070

Patient Name : EQAS 1 CP DEC23

Age / Gender : 1 years / Male

Mobile No. : -

Patient ID : 3773

Source : DIRECT

Scan to Validate



Referral : SELF

Collection Time : Dec 01, 2023, 03:52 p.m.

Receiving Time : Dec 01, 2023, 03:52 p.m.

Reporting Time : Dec 12, 2023, 03:39 p.m.

Sample ID :



002833523

Test Description	Value(s)	Reference Range	Unit(s)
URINE EXAMINATION, ROUTINE; URINE, R/E			
Volume*	4	-	ml
Colour*	Pale Yellow	Pale Yellow	
Transparency (Appearance)*	Clear	Clear	
Deposit*	Absent	Absent	
Reaction (pH)*	6.5	4.5 - 8	
Specific Gravity*	1.025	1.010 - 1.030	
Chemical Examination (Automated Dipstick Method) Urine			
Urine Glucose (sugar)*	Present (+)	Absent	
Urine Protein (Albumin)*	Present (++)	Absent	
Urine Ketones (Acetone)*	Absent	Absent	
Blood*	Present(+++)	Absent	
Bile pigments*	Absent	Absent	
Nitrite*	Present	Absent	
Urobilinogen*	Normal	Normal	
Microscopic Examination Urine			
Pus Cells (WBCs)*	0-1	0 - 5	/hpf
Epithelial Cells*	0-1	0 - 4	/hpf
Red blood Cells*	Absent	Absent	/hpf
Crystals*	Absent	Absent	
Cast*	Absent	Absent	
Trichomonas Vaginalis*	Absent	Absent	
Yeast Cells*	Absent	Absent	
Amorphous deposits*	Absent	Absent	
Bacteria*	Absent	Absent	

END OF REPORT

Reported By : Aniket Shinde

Dr. Supriya Lad Chincholkar
MD (PATH), DNB (PATH), DPB
MMC Reg 076789