



**METSAR TECHNOLOGIES PVT. LTD.**  
Calibration Laboratory



**CALIBRATION CERTIFICATE**

<b>Customer Name &amp; Address</b>	<b>Certificate No.</b> : MTPL/23/1957/4
<b>EMPLOYEE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL,</b> Sanathnagar, Hyderabad – 500038.	<b>Equipment Received On</b> : 06/10/2023
	<b>Equipment Condition</b> : Satisfactory
	<b>Date of Calibration</b> : 07/10/2023
	<b>Recommended Calibration Due</b> : 06/10/2024
	<b>Date of issue</b> : 10/10/2023
<b>MECHANICAL DISCIPLINE (VOLUME)</b>	<b>ULR No.</b> : CC219123000019203F

**Details of Unit Under Calibration:**

<b>Instrument Specification</b> : Micro Pipette	<b>Range</b> : 2 to 20 µl
<b>Make</b> : Eppendorf	<b>Resolution</b> : 0.02 µl
<b>Model</b> : Research Plus	<b>Unit Under Measurement</b> : µl
<b>Sr. No</b> : O27153F	

**Standard used for calibration:**

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Micro Balance	METSAR-M-001	MTPL/23/0156/1	26/01/2024	CC-2191

Environmental Conditions		SOP Number
<b>Temperature</b> 23 ± 1 °C	<b>Humidity</b> 40 to 60 % RH	MTPL/CL/SOP/MV/03

**Results of Calibration:**

Sr. No.	Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume in µl @ 27 °C	Error (±) in µl	Expanded Uncertainty (±) in µl
1	2.00	0.002019	2.014	0.014	0.1
2	5.00	0.005035	5.021	0.021	0.1
3	10.00	0.010062	10.035	0.035	0.1
4	15.00	0.015079	15.039	0.039	1.2
5	20.00	0.020095	20.043	0.043	1.2

**Repeatability Results @ 27 °C:**

Sr. No.	Set Volume in µl	Actual Calculated Volume in µl				
1	20.00	20.047	20.041	20.038	20.040	20.043
		20.044	20.047	20.043	20.044	20.040

**Remarks:**

- The Standard used for calibration is traceable to National/International standards through unbroken chain of Accredited Laboratories.
- Reference Standard and Method used: ISO 8655-6, Gravimetric Method.
- The result stated in this calibration certificate is related only to the item submitted for calibration.
- UUC means Unit Under Calibration.
- The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor  $k=2$ .
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The instrument was calibrated at Mass and Volume Lab.
- The Recommended Due Date of this calibration certificate is given as per request of customer.

*CH. Naresh*  
Calibrated By  
**CH. Naresh**  
Sr. Calibration Engineer

*M. Ramratan*  
Certificate Approved By  
**M. Ramratan**  
Technical Manager

**\*\*End of Calibration Certificate\*\***

**METSAR TECHNOLOGIES PVT. LTD.**

An ISO 9001-2015 Certified Company





**METSAR TECHNOLOGIES PVT. LTD.**  
Calibration Laboratory



**CALIBRATION CERTIFICATE**

<b>Customer Name &amp; Address</b>	<b>Certificate No.</b> : MTPL/23/1957/2
<b>EMPLOYEE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL,</b> Sanathnagar, Hyderabad – 500038.	<b>Equipment Received On</b> : 06/10/2023
	<b>Equipment Condition</b> : Satisfactory
	<b>Date of Calibration</b> : 07/10/2023
	<b>Recommended Calibration Due</b> : 06/10/2024
	<b>Date of issue</b> : 10/10/2023
<b>MECHANICAL DISCIPLINE (VOLUME)</b>	<b>ULR No.</b> : CC219123000019201F

**Details of Unit Under Calibration:**

<b>Instrument Specification</b> : Micro Pipette	
<b>Make</b> : Eppendorf	
<b>Model</b> : Research Plus	<b>Range</b> : 100 to 1000 µl
<b>Sr. No</b> : O22503F	<b>Resolution</b> : 1 µl
<b>Id. No.</b> : ---	<b>Unit Under Measurement</b> : g

**Standard used for calibration:**

<b>Instrument Name</b>	<b>Instrument Sr. No. / Id No.</b>	<b>Certificate No.</b>	<b>Calibration Due On</b>	<b>Traceability with NABL Lab No.</b>
Semi Micro Balance	METSAR-M-002	MTPL/23/0156/2	26/01/2024	CC-2191

<b>Environmental Conditions</b>		<b>SOP Number</b>
<b>Temperature</b>	<b>Humidity</b>	MTPL/CL/SOP/MV/03
23 ± 1 °C	40 to 60 % RH	

**Results of Calibration:**

<b>Sr. No.</b>	<b>Set Volume in µl</b>	<b>Standard Balance Reading in g</b>	<b>Actual Calculated Volume in µl @ 27 °C</b>	<b>Error (±) in µl</b>	<b>Expanded Uncertainty (±) in µl</b>
1	100	0.10037	100.09	0.09	15
2	300	0.30095	300.11	0.11	15
3	500	0.50157	500.21	0.21	15
4	800	0.80246	800.27	0.27	15
5	1000	1.00315	1000.39	0.39	15

**Repeatability Results @ 27 °C :**

<b>Sr. No.</b>	<b>Set Volume In µl</b>	<b>Actual Calculated Volume In µl</b>				
1	1000	1000.39	1000.39	1000.43	1000.39	1000.36
		1000.42	1000.38	1000.35	1000.43	1000.39

**Remarks:**

- The Standard used for calibration is traceable to National/International standards through unbroken chain of Accredited Laboratories.
- Reference Standard and Method used: ISO 8655-6, Gravimetric Method.
- The result stated in this calibration certificate is related only to the item submitted for calibration.
- UUC means Unit Under Calibration.
- The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor  $k=2$ .
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The instrument was calibrated at Mass and Volume Lab.
- The Recommended Due Date of this calibration certificate is given as per request of customer.

Ch. Naresh

Calibrated By  
CH. Naresh  
Sr. Calibration Engineer

Certificate Approved By  
M. Ramratan  
Technical Manager

**\*\*End of Calibration Certificate\*\***

**METSAR TECHNOLOGIES PVT. LTD.**

An ISO 9001-2015 Certified Company

MTPL/CL/FF/CC/ME/MP

Page 1 of 1





**METSAR TECHNOLOGIES PVT. LTD.**  
Calibration Laboratory



**CALIBRATION CERTIFICATE**

<b>Customer Name &amp; Address</b>	<b>Certificate No.</b> : MTPL/23/1957/1
<b>EMPLOYEE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL,</b> Sanathnagar, Hyderabad – 500038.	<b>Equipment Received On</b> : 06/10/2023
	<b>Equipment Condition</b> : Satisfactory
	<b>Date of Calibration</b> : 07/10/2023
	<b>Recommended Calibration Due</b> : 06/10/2024
	<b>Date of issue</b> : 10/10/2023
<b>MECHANICAL DISCIPLINE (VOLUME)</b>	<b>ULR No.</b> : CC219123000019200F

**Details of Unit Under Calibration:**

<b>Instrument Specification</b> : Micro Pipette	
<b>Make</b> : Eppendorf	
<b>Model</b> : Research Plus	<b>Range</b> : 20 to 200 µl
<b>Sr. No</b> : P31472F	<b>Resolution</b> : 0.2 µl
<b>Id. No.</b> : ---	<b>Unit Under Measurement</b> : g

**Standard used for calibration:**

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Semi Micro Balance	METSAR-M-002	MTPL/23/0156/2	26/01/2024	CC-2191
Micro Balance	METSAR-M-001	MTPL/23/0156/1	26/01/2024	CC-2191

Environmental Conditions		SOP Number
<b>Temperature</b> 23 ± 1°C	<b>Humidity</b> 40 to 60 %RH	MTPL/CL/SOP/MV/03

**Results of Calibration:**

Sr. No.	Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume in µl @ 27 °C	Error (±) in µl	Expanded Uncertainty (±) in µl
1	20.0	0.020030	19.989	0.011	1.2
2	50.0	0.05022	50.12	0.12	1.2
3	100.0	0.10044	100.23	0.23	1.2
4	150.0	0.15062	150.31	0.31	15
5	200.0	0.20074	200.38	0.38	15

**Repeatability Results @ 27 °C :**

Sr. No.	Set Volume In µl	Actual Calculated Volume In µl				
		1	200.0	200.38	200.34	200.36
		200.40	200.43	200.34	200.45	200.34

**Remarks:**

- The Standard used for calibration is traceable to National/International standards through unbroken chain of Accredited Laboratories.
- Reference Standard and Method used: ISO 8655-6, Gravimetric Method.
- The result stated in this calibration certificate is related only to the item submitted for calibration.
- UUC means Unit Under Calibration.
- The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor  $k=2$ .
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The instrument was calibrated at Mass and Volume Lab.
- The Recommended Due Date of this calibration certificate is given as per request of customer.

Ch. Naresh  
Calibrated By  
CH. Naresh  
Sr. Calibration Engineer

Certificate Approved By  
M. Ramratan  
Technical Manager

**METSAR TECHNOLOGIES PVT. LTD.**  
\*\*End of Calibration Certificate\*\*  
An ISO 9001-2015 Certified Company





**METSAR TECHNOLOGIES PVT. LTD.**  
Calibration Laboratory

**CALIBRATION CERTIFICATE**



CC-2191

<b>Customer Name &amp; Address</b>	<b>Certificate No.</b> : MTPL/23/1607/19
<b>EMPLOYEE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL,</b> Sanathnagar, Hyderabad – 500038.	<b>Equipment Received On</b> : 19/08/2023
	<b>Equipment Condition</b> : Satisfactory
	<b>Date of Calibration</b> : 19/08/2023
	<b>Recommended Calibration Due</b> : 18/08/2024
	<b>Date of issue</b> : 21/08/2023
<b>MECHANICAL DISCIPLINE (VOLUME)</b>	<b>ULR No.</b> : CC219123000018554F

**Details of Unit Under Calibration:**

<b>Instrument Specification</b> : Micro Pipette	<b>Range</b> : 2 to 20 µl
<b>Make</b> : Eppendorf	<b>Resolution</b> : 0.02 µl
<b>Model</b> : Research Plus	<b>Unit Under Measurement</b> : µl
<b>Sr. No</b> : Q27204F	

**Standard used for calibration:**

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Micro Balance	METSAR-M-001	MTPL/23/0156/1	26/01/2024	CC-2191

Environmental Conditions		SOP Number
<b>Temperature</b> 23 ± 1 °C	<b>Humidity</b> 40 to 60 % RH	MTPL/CL/SOP/MV/03

**Results of Calibration:**

Sr. No.	Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume in µl @ 27 °C	Error (±) in µl	Expanded Uncertainty (±) in µl
1	2.00	0.002019	2.014	0.014	0.1
2	5.00	0.005010	4.995	0.005	0.1
3	10.00	0.010039	10.008	0.008	0.1
4	15.00	0.015021	14.976	0.024	1.2
5	20.00	0.019987	19.928	0.072	1.2

**Repeatability Results @ 27 °C:**

Sr. No.	Set Volume in µl	Actual Calculated Volume in µl				
1	20.00	19.927	19.925	19.933	19.923	19.928
		19.923	19.935	19.929	19.925	19.931

**Remarks:**

- The Standard used for calibration is traceable to National/International standards through unbroken chain of Accredited Laboratories.
- Reference Standard and Method used: ISO 8655-6, Gravimetric Method.
- The result stated in this calibration certificate is related only to the item submitted for calibration.
- UUC means Unit Under Calibration.
- The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor  $k=2$ .
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The instrument was calibrated at Mass and Volume Lab-1.

*Ch. Naresh*

Calibrated By

CH. Naresh

Sr. Calibration Engineer

*N. Chanakya*

Certificate Approved By

N. Chanakya

Sr. Manager Calibration

**\*\*End of Calibration Certificate\*\***

**METSAR TECHNOLOGIES PVT. LTD.**

An ISO 9001-2015 Certified Company





# METSAR TECHNOLOGIES PVT. LTD.

## Calibration Laboratory



### CALIBRATION CERTIFICATE

<b>Customer Name &amp; Address</b>	<b>Certificate No.</b> : MTPL/23/1607/20
<b>EMPLOYE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL,</b> Sanathnagar, Hyderabad – 500038.	<b>Equipment Received On</b> : 19/08/2023
	<b>Equipment Condition</b> : Satisfactory
	<b>Date of Calibration</b> : 19/08/2023
	<b>Recommended Calibration Due</b> : 18/08/2024
	<b>Date of issue</b> : 21/08/2023
<b>MECHANICAL DISCIPLINE (VOLUME)</b>	<b>ULR No.</b> : CC219123000018555F

#### Details of Unit Under Calibration:

<b>Instrument Specification</b> : Micro Pipette	<b>Range</b> : 0.5 to 10 µl
<b>Make</b> : Eppendorf	<b>Resolution</b> : 0.01 µl
<b>Model</b> : Research	<b>Unit Under Measurement</b> : g
<b>Sr. No</b> : 400144Z	

#### Standard used for calibration:

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Micro Balance	METSAR-M-001	MTPL/23/0156/1	26/01/2024	CC-2191

Environmental Conditions		SOP Number
<b>Temperature</b> 23 ± 1 °C	<b>Humidity</b> 40 to 60 % RH	MTPL/CL/SOP/MV/03

#### Results of Calibration:

Sr. No.	Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume in µl @ 27 °C	Error (+) in µl	Expanded Uncertainty (+) in µl
1	1.00	0.001008	1.005	0.005	0.1
2	3.00	0.003023	3.015	0.015	0.1
3	5.00	0.005031	5.016	0.016	0.1
4	8.00	0.008068	8.047	0.047	0.1
5	10.00	0.010084	10.057	0.057	0.1

#### Repeatability Results @ 27 °C:

Sr. No.	Set Volume in µl	Actual Calculated Volume in µl				
		1	10.00	10.058	10.052	10.057
		10.060	10.067	10.060	10.048	10.057

#### Remarks:

- The Standard used for calibration is traceable to National/International standards through unbroken chain of Accredited Laboratories.
- Reference Standard and Method used: ISO 8655-6, Gravimetric Method.
- The result stated in this calibration certificate is related only to the item submitted for calibration.
- UUC means Unit Under Calibration.
- The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor  $k=2$ .
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The instrument was calibrated at Mass and Volume Lab-1.

*Ch. Naresh*

Calibrated By

CH. Naresh

Sr. Calibration Engineer

*N. Chanakya*

Certificate Approved By

N. Chanakya

Sr. Manager Calibration

\*\* End of Calibration Certificate\*\*



METSAR TECHNOLOGIES PVT. LTD.

An ISO 9001-2015 Certified Company

1st Floor, Garg Trade Centre, Balanagar, Medchal - Malkajgiri Dist., Hyderabad - 500 037, Telangana, INDIA. © +91 9640166643 / +91-40-29800836, info@metsartechnologies.com

Page 1 of 1

www.metsartechnologies.com





# METSAR TECHNOLOGIES PVT. LTD.

Calibration Laboratory



## CALIBRATION CERTIFICATE

<b>Customer Name &amp; Address</b>	<b>Certificate No.</b> : MTPL/23/1607/12
<b>EMPLOYE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL,</b> Sanathnagar, Hyderabad – 500038.	<b>Equipment Received On</b> : 19/08/2023
	<b>Equipment Condition</b> : Satisfactory
	<b>Date of Calibration</b> : 19/08/2023
	<b>Recommended Calibration Due</b> : 18/08/2024
	<b>Date of issue</b> : 21/08/2023
<b>MECHANICAL DISCIPLINE (VOLUME)</b>	<b>ULR No.</b> : CC219123000018547F

### Details of Unit Under Calibration:

<b>Instrument Specification</b> : Micro Pipette			
<b>Make</b> : Eppendorf	<b>Range</b> : 100 to 1000 µl		
<b>Model</b> : Research	<b>Resolution</b> : 1 µl		
<b>Sr. No</b> : 486345Z	<b>Unit Under Measurement</b> : µl		

### Standard used for calibration:

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Semi Micro Balance	METSAR-M-002	MTPL/23/0156/2	26/01/2024	CC-2191

Environmental Conditions		SOP Number
<b>Temperature</b> 23 ± 1°C	<b>Humidity</b> 40 to 60 %RH	MTPL/CL/SOP/MV/03

### Results of Calibration:

Sr. No.	Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume in µl @ 27 °C	Error (+) in µl	Expanded Uncertainty (+) in µl
1	100	0.10024	100.03	0.03	1.2
2	300	0.30019	299.52	0.48	15
3	500	0.50056	499.45	0.55	15
4	800	0.80088	799.11	0.89	15
5	999	0.99868	996.58	2.42	15

### Repeatability Results @ 27 °C:

Sr. No.	Set Volume In µl	Actual Calculated Volume In µl				
1	999	996.54	996.59	996.55	996.53	996.62
		996.65	996.69	996.58	996.51	996.55

### Remarks:

- The Standard used for calibration is traceable to National/International standards through unbroken chain of Accredited Laboratories.
- Reference Standard and Method used: ISO 8655-6, Gravimetric Method.
- The result stated in this calibration certificate is related only to the item submitted for calibration.
- UUC means Unit Under Calibration.
- The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor  $k=2$ .
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The instrument was calibrated at Mass and Volume Lab-1.

*Ch. Naresh*  
Calibrated By

CH. Naresh  
Sr. Calibration Engineer

*N. Chanakya*  
Certificate Approved By  
N. Chanakya  
Sr. Manager Calibration

\*\*End of Calibration Certificate\*\*

MTPL/CL/FF/CC/ME/MP

METSAR TECHNOLOGIES PVT. LTD.

An ISO 9001-2015 Certified Company

Page 1 of 1





# METSAR TECHNOLOGIES PVT. LTD.

## Calibration Laboratory



### CALIBRATION CERTIFICATE

<b>Customer Name &amp; Address</b>	<b>Certificate No.</b> : MTPL/23/1607/14
<b>EMPLOYE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL,</b> Sanathnagar, Hyderabad – 500038.	<b>Equipment Received On</b> : 19/08/2023
	<b>Equipment Condition</b> : Satisfactory
	<b>Date of Calibration</b> : 19/08/2023
	<b>Recommended Calibration Due</b> : 18/08/2024
	<b>Date of issue</b> : 21/08/2023
<b>MECHANICAL DISCIPLINE (VOLUME)</b>	<b>ULR No.</b> : CC219123000018549F

#### Details of Unit Under Calibration:

<b>Instrument Specification</b> : Micro Pipette		
<b>Make</b> : Eppendorf	<b>Range</b> : 100 to 1000 µl	
<b>Model</b> : Research Plus	<b>Resolution</b> : 1 µl	
<b>Sr. No</b> : O22291F	<b>Unit Under Measurement</b> : µl	

#### Standard used for calibration:

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Semi Micro Balance	METSAR-M-002	MTPL/23/0156/2	26/01/2024	CC-2191

Environmental Conditions		SOP Number
<b>Temperature</b> 23 ± 1°C	<b>Humidity</b> 40 to 60 %RH	MTPL/CL/SOP/MV/03

#### Results of Calibration:

Sr. No.	Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume in µl @ 27 °C	Error (±) in µl	Expanded Uncertainty (±) in µl
1	100	0.10027	100.05	0.05	1.2
2	300	0.30040	299.73	0.27	15
3	500	0.60040	599.08	0.92	15
4	800	0.80060	798.82	1.18	15
5	1000	1.00080	998.71	1.29	15

#### Repeatability Results @ 27 °C :

Sr. No.	Set Volume in µl	Actual Calculated Volume in µl			
1	1000	998.69	998.77	998.73	998.66
		998.77	998.73	998.77	998.65

#### Remarks:

- The Standard used for calibration is traceable to National/International standards through unbroken chain of Accredited Laboratories.
- Reference Standard and Method used: ISO 8655-6, Gravimetric Method.
- The result stated in this calibration certificate is related only to the item submitted for calibration.
- UUC means Unit Under Calibration.
- The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor  $k=2$ .
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The instrument was calibrated at Mass and Volume Lab-1.

eh.nareesh.

Calibrated By  
CH. Naresh

Sr. Calibration Engineer

Certificate Approved By  
N. Chanakya  
Sr. Manager Calibration

\*\*End of Calibration Certificate\*\*

MTPL/CL/FF/CC/ME/MP

METSAR TECHNOLOGIES PVT. LTD.

Page 1 of 1





# METSAR TECHNOLOGIES PVT. LTD.

## Calibration Laboratory



### CALIBRATION CERTIFICATE

<b>Customer Name &amp; Address</b>	<b>Certificate No.</b> : MTPL/23/1607/18
<b>EMPLOYEE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL,</b> Sanathnagar, Hyderabad – 500038.	<b>Equipment Received On</b> : 19/08/2023
	<b>Equipment Condition</b> : Satisfactory
	<b>Date of Calibration</b> : 19/08/2023
	<b>Recommended Calibration Due</b> : 18/08/2024
	<b>Date of issue</b> : 21/08/2023
<b>MECHANICAL DISCIPLINE (VOLUME)</b>	<b>ULR No.</b> : CC219123000018553F

#### Details of Unit Under Calibration:

<b>Instrument Specification</b> : Micro Pipette	
<b>Make</b> : Eppendorf	<b>Range</b> : 500 - 5000 µl
<b>Model</b> : Research	<b>Resolution</b> : 5 µl
<b>Sr. No</b> : 498459Z	<b>Unit Under Measurement</b> : µl

#### Standard used for calibration:

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Semi Micro Balance	METSAR-M-002	MTPL/23/0156/2	26/01/2024	CC-2191

Environmental Conditions		SOP Number
Temperature	Humidity	MTPL/CL/SOP/MV/03
23 ± 1°C	40 to 60 % RH	

#### Results of Calibration:

Sr. No.	Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume in µl @ 27 °C	Error (±) in µl	Expanded Uncertainty (±) in µl
1	500	0.49986	498.81	1.19	15
2	2000	1.99859	1994.16	5.84	15
3	3000	2.99632	2989.68	10.32	15
4	4000	3.99458	3985.71	14.29	15
5	5000	4.99312	4982.60	17.40	15

#### Repeatability Results @ 27 °C:

Sr. No.	Set Volume In µl	Actual Calculated Volume In µl				
1	5000	4982.60	4982.64	4982.56	4982.56	4982.54
		4982.58	4982.63	4982.60	4982.68	4982.65

#### Remarks:

- The Standard used for calibration is traceable to National/International standards through unbroken chain of Accredited Laboratories.
- Reference Standard and Method used: ISO 8655-6, Gravimetric Method.
- The result stated in this calibration certificate is related only to the item submitted for calibration.
- UUC means Unit Under Calibration.
- The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor  $k=2$ .
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The instrument was calibrated at Mass and Volume Lab-1.

*eh. Naresh.*  
Calibrated By  
CH. Naresh  
Sr. Calibration Engineer

*[Signature]*  
Certificate Approved By  
N. Chanakya  
Sr. Manager Calibration

\*\* End of Calibration Certificate \*\*

**METSAR TECHNOLOGIES PVT. LTD.**

An ISO 9001-2015 Certified Company





**METSAR TECHNOLOGIES PVT. LTD.**  
Calibration Laboratory



**CALIBRATION CERTIFICATE**

<b>Customer Name &amp; Address</b>	<b>Certificate No.</b> : MTPL/23/1957/3
<b>EMPLOYEE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL,</b> Sanathnagar, Hyderabad – 500038.	<b>Equipment Received On</b> : 06/10/2023
	<b>Equipment Condition</b> : Satisfactory
	<b>Date of Calibration</b> : 07/10/2023
	<b>Recommended Calibration Due</b> : 06/10/2024
	<b>Date of issue</b> : 10/10/2023
<b>MECHANICAL DISCIPLINE (VOLUME)</b>	<b>ULR No.</b> : CC219123000019202F

**Details of Unit Under Calibration:**

<b>Instrument Specification</b> : Micro Pipette	
<b>Make</b> : Erba	
<b>Model</b> : ---	<b>Range</b> : 5 - 50 µl
<b>Sr. No</b> : AB08492	<b>Resolution</b> : 0.5 µl
<b>Id. No.</b> : ---	<b>Unit Under Measurement</b> : g

**Standard used for calibration:**

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Semi Micro Balance	METSAR-M-002	MTPL/23/0156/2	26/01/2024	CC-2191
Micro Balance	METSAR-M-001	MTPL/23/0156/1	26/01/2024	CC-2191

Environmental Conditions		SOP Number
<b>Temperature</b> 23 ± 1°C	<b>Humidity</b> 40 to 60 %RH	MTPL/CL/SOP/MV/03

**Results of Calibration:**

Sr. No.	Set Volume in µl	Standard Balance Reading in g	Actual Calculated Volume in µl @ 27 °C	Error (±) in µl	Expanded Uncertainty (±) in µl
1	5.0	0.005021	5.010	0.010	1.2
2	20.0	0.02015	20.11	0.11	1.2
3	30.0	0.03019	30.14	0.14	1.2
4	40.0	0.04036	40.29	0.29	1.2
5	50.0	0.05046	50.37	0.37	1.2

**Repeatability Results @ 27 °C :**

Sr. No.	Set Volume in µl	Actual Calculated Volume in g				
		1	50.0	50.33	50.41	50.38
		50.35	50.40	50.34	50.38	50.36

**Remarks:**

- The Standard used for calibration is traceable to National/International standards through unbroken chain of Accredited Laboratories.
- Reference Standard and Method used: ISO 8655-6, Gravimetric Method.
- The result stated in this calibration certificate is related only to the item submitted for calibration.
- UUC means Unit Under Calibration.
- The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor  $k=2$ .
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The instrument was calibrated at Mass and Volume Lab.
- The Recommended Due Date of this calibration certificate is given as per request of customer.

Ch. Naresh

Calibrated By  
**CH. Naresh**  
Sr. Calibration Engineer

Certificate Approved By  
**M. Ramratan**  
Technical Manager

**METSAR TECHNOLOGIES PVT. LTD.**

**\*End of Calibration Certificate\*\***  
An ISO 9001-2015 Certified Company