



METSAR TECHNOLOGIES PVT. LTD.

Calibration Laboratory

CALIBRATION CERTIFICATE



Customer Name & Address	Certificate No. : MTPL/23/1619/1
EMPLOYEE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL, Sanathnagar, Hyderabad – 500038.	Equipment Received On : 21/08/2023
	Equipment Condition : Satisfactory
	Date of Calibration : 21/08/2023 ✓
	Recommended Calibration Due : 20/08/2024 ✓
	Date of Issue : 22/08/2023
MECHANICAL DISCIPLINE (MASS)	ULR No. : CC219123000017963F

Details of Unit Under Calibration:

Instrument Specification : Weighing Balance			
Make : WENSAR	Location : Microbiology		
Model : PGB 200	Capacity : 200g		
Sr. No. : 16029	Readability : 0.001 g		
Id. No. : ESICSSH/MICRO/EQ 6	Unit Under Measurement : g		

Standard used for calibration:

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Set of Weights	METSAR-M-017	MTPL/23/0526/6	24/03/2024	CC-2191


Environmental Conditions

Temperature
25 ± 3 °C

Humidity
30 to 75 %RH

Calibration Method: Balance Calibration Procedure (MTPL/CL/SOP/MV/01)

Reference Standard: OIML R 76: 2006


Calibrated By
CH. Sairam
Calibration Engineer


Certificate Approved By
M. Ramratan
Technical Manager

MTPL/CL/FF/CC/ME/WB

METSAR TECHNOLOGIES PVT. LTD.

Page 1 of 3