

# METSAR TECHNOLOGIES PVT. LTD.

## Calibration Laboratory







Customer Name & Address	Certificate No.	: MTPL/23/1619/3
EMPLOYE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL, Sanathnagar, Hyderabad – 500038.	Equipment Received On	: 21/08/2023
	Equipment Condition	: Satisfactory
	Date of Calibration	: 21/08/2023
	Recommended Calibration Due	: 20/08/2024
	Date of Issue	: 22/08/2023
MECHANICAL(SPEED)DISCIPLINE	ULR No:	: CC219123000017965F

#### Details of Unit Under Calibration:

Instrument Specification	: Centrifuge	A SA CALL	A Second Second
Make	: REMI	Location	: Microbiology
Model	: R-8C BL	Range	: 0 to 6000 rpm
Id No.	: LEC-13266	Resolution	: 10 rpm
Location	11-	Unit Under Measurement	: rpm

### Standard used for calibration:

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On	Traceability with NABL Lab No.
Digital Tachometer	METSAR-TM-005	CC/ECL/1665/22-23	22/01/2024	CC-2287

Environmen	ntal Conditions	SOP Number	
Temperature	Humidity	MTDL/CL/COD/ME/01	
$23 \pm 3^{\circ}\text{C}$	40 to 60% RH	MTPL/CL/SOP/ME/01	

### Results of Calibration:

Sr. No.	Set Value in rpm	STD Reading in rpm	UUC Reading in rpm	Error (±) in rpm	Expanded Uncertainty (±) in rpm
1	100	100.02 "	100	0.02	3.3
2	500	500.05	500	0.05	3.3
3	1000	1000.1	1000	0.1	3.3
4	3000	3000.2	3000	0.2	3.3
5	6000	6000.5	6000	0.5	3.3

#### Remarks:

- a) The Standard used for calibration traceable to National/International standards through unbroken chain of Accredited Laboratories.
- b) The results stated in this certificate relate only to the item calibrated.
- c) UUC means Unit Under Calibration.
- d) The reported expanded uncertainty of measurement is stated at a confidence level of approximately 95.45% with coverage factor k=2.
- e) Certificate shall not be reproduced except in full without the written approval of the laboratory.
- f) Results reported are valid at the time of and under the stated conditions of measurements.
- g) The above UUC was calibrated at Site.

Calibrated By
CH. Sairam
Calibration Engineer

Certificate Approved By
M. Ramratan
Technical Manager

\*\*End of Calibration Certificate\*\*

MTPL/CL/FF/CC/ME/RM

METSAR TECHNOLOGIES PVT. LTD.
An ISO 9001-2015 Certified Company

Page 1 of 1