



METSAR TECHNOLOGIES PRIVATE LIMITED

(An ISO 9001 : 2015 Certified Company)

Calibration Laboratory

Calibration Certificate

Customer Name & Address EMPLOYEE'S STATE INSURANCE CORPORATION SUPER SPECIALITY HOSPITAL, Sanathnagar, Hyderabad – 500038.	Certificate No. : MTPL/23/1619/NN/4 Equipment Received On : 21/08/2023 Equipment Condition : Satisfactory Date of Calibration : 21/08/2023 Recommended Calibration Due : 20/08/2024 Date of Issue : 22/08/2023
---	---

Details of Unit Under Calibration:

Instrument Specification : Cyclo mixer Make : REMI Model : CM 101 Sr. No. : EECM698 Id. No. : ESICSSH/MICRO/EQ45	Location : Microbiology Range : 2500 rpm Unit Under Measurement : rpm
---	--

Standard used for calibration:

Instrument Name	Instrument Sr. No. / Id No.	Certificate No.	Calibration Due On
Digital Tachometer	METSAR-TM-005	CC/ECL/1665/22-23	22/01/2024

Environmental Conditions		SOP Number
Temperature 23 ± 3°C	Humidity 40 to 60% RH	MTPL/CL/SOP/ME/01

Results of Calibration:

Sr. No.	Set Value in rpm	STD Reading in rpm	Error (±) in rpm
1	2500	2502.4	2.4

Remarks:

- The Standard used for calibration traceable to National/International standards through unbroken chain of Accredited Laboratories.
- The results stated in this certificate relate only to the item calibrated.
- UUC means Unit under calibration.
- Certificate shall not be reproduced except in full without the written approval of the laboratory.
- The above UUC was calibrated at site.


Calibrated By
CH. Sairam
Calibration Engineer


Certificate Approved By
M. Ramratan
Technical Manager

** End of Calibration Certificate**

MTPL/CL/FF/CC/ME/RM

Page 1 of 1

METSAR TECHNOLOGIES PVT. LTD