



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 09160F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	231215.6	DT - 15 / 12 / 2023		ULR No.	CC2646230000 09160F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	16.12.2023		
Address	DH Civil Hospital Guru Teg Bahadur Nagar Tarn Taran			Recommended Date for Next Calib.	15.12.2024		
				Certificate Issue Dt.	18.12.2023		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	100 to 1000 µl	5 µl	Dragon Lab YE225AX0315476	KDTPBMLMP08 Main Lab

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6 - 2022	Volume Measurement	55 %RH	23.7 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in µl)
1	100	100.00269	-0.00269	0.25
2	500	500.00541	-0.00541	0.25
3	1000	1000.00935	-0.00935	0.25

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Asluch
Calibrated by



Approved by
Asluch
Anand Chaurasia
(CEO)