



# The Accurate Calibration & Medical Solution CALIBRATION CERTIFICATE

Page 1 of 1

## CALIBRATION CERTIFICATE OF MICROSCOPE

Certificate No :	TAC/12-23/KDL/MS-1001	Callbration Date :	16.12.2023
Job Identification No :	TAC/C/16/12/2023	Next Callbration Date :	15.12.2024
Ref. No :	SRF Dated - 14. 12. 2023	Issue Date :	10.12.2023

Customer Name	M/S KRSNAA DIAGNOSTICS LTD.
Address	DH CIVIL HOSPITAL GURU TEG BAHADUR NAGAR TARN TARAN -143401

## DESCRIPTION OF THE EQUIPMENT TO BE CALIBRATED

Instrument Name	MICRO SCOPE		
Make	LABOMED	Visual Inspection	Ok
Model	-----	I.D. No.	KDTPBCPMS05
Serial No.	230147067	Location / Dept.	CLINICAL PATHOLOGY

## CALIBRATION STANDARD USED

Sl.No.	Master Equipment	Manufacturer	Model No.	Serial No.	Date of Calibration	Valld Upto
1	Electrical Safety Analyzer	DATREND Systems Inc.	vPed - Mini	VPM18100028	14.10.2023	14.10.2024

ENVIRONMENTAL CONDITIONS	Temperature	22.0 °C ± 2.0 °C	Humidity	50 ± 10%RH
--------------------------	-------------	------------------	----------	------------

## RESULT

### VISUAL TEST

Sr.No.	Tests	Remarks
1	Power Cords,Cable Checking	OK
2	Mains Socket Checking	OK
3	Equipment Type (B,BF,CF)	CF
4	Equipment Class ( I, II, III )	I


### FUNCTIONAL TEST

Sr.No.	Parameters	Measured Value	Remarks
1	Voltage Between Live and Neutral (VLN)	231.6 V (220-240 V)	OK
2	Voltage Between Live and Earth (VLE)	230.8 V (220-240V)	OK
3	Voltage Between Neutral and Earth (VNE)	0.3 V (0-5V)	OK
4	Load Current (Ia)	0.7 A	OK
5	Enclosure Current	0.8 µA (0-10µA)	OK
6	Leaking Current (IL)	102 µA (0-300µA)	OK
7	Time for Testing (t)	15 (min)	OK


Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed

YES



NO




Remarks:

- 1.Certificates shall not be reproduced except the written approval of The Accurate Calibration & Medical Solution.
- 2.The Calibration result reported in this certificate are only valid for the time of calibration and the stated condition of measurement.

Calibrated by

\*\*\* END OF CERTIFICATE \*\*\*



Approved by

Plot No. 72/G1, Ganesh Vihar Sirsi Road, Jaipur (Raj.) -302012

Mob.: 9672294758

E-mail : theaccuratecalibration@gmai.com