



**SD BIOSENSOR HEALTHCARE PVT. LTD.**

Registered Office cum R&D Center : Plot No. 63, Udyog Vihar, Phase-IV, Gurugram-122015, Haryana, India  
Tel.:(+91) 124-454-0907, Toll Free No : (+91) 1800 10 23105 Website: www.sdbiosensor.co.in : Email: care@sdbiosensor.co.in  
CIN - U33110HR2012PTC047034

---

**Installation Certificate for Standard F 200**

This is to certify that Standard F 200 instrument Serial No. FA20FAITG1461 is successfully installed and commissioned at Lupin Diagnostic, Jalna and the installation protocol/checklist has been successfully completed for the above instrument.

For **SD Biosensor Healthcare Pvt. Ltd.**

A handwritten signature in black ink, appearing to read "Aamir".



Aamir Ayaz Khan

Assistant Product Manager

# SD BIOSENSOR

SD BIOSENSOR HEALTHCARE PVT. LTD.

Registered Office cum R&D Center : Plot No. 63, Udyog Vihar, Phase-IV, Gurugram-122015, Haryana, India  
Tel.:(+91) 124-454-0907, Toll Free No : (+91) 1800 10 23105 Website: www.sdbiosensor.co.in : Email: care@sdbiosensor.co.in  
CIN - U33110HR2012PTC047034

## Installation Report

Customer Details	Instrument Details
------------------	--------------------

Name: Lupin Diagnostic	Serial Number: FA20FAITG1461
Model: Standard F 200	
Address: D7 & D8, Mahalaxmi Nagar, Mantha Chowfully, Jalna 431203	
Contact Person(s): Ajay Wayal	Product of: SD Biosensor
Contact No: 8888216843	

Accessories and Consumable Supplied
-------------------------------------

1. Power adapter: Yes
2. Calibration set: Yes
3. Test kit, Buffer: Yes
4. User instruction guide: Yes


Installation Checks
---------------------

1. Environment: Closed
2. Power Adapter: Working
3. Temperature: 26 degree Celsius (in limits between 15 degree to 32 degree Celsius)
4. UPS: Yes


Training Imparted	Assays Performed
-------------------	------------------


1. Operating Procedure
2. Maintenance
3. Dos and Don'ts
4. Calibration

Customer Remarks	Installer's Remarks
------------------	---------------------



Customer's Signature





Installer's Signature



**SD BIOSENSOR HEALTHCARE PVT. LTD.**


Registered Office cum R&D Center : Plot No. 63, Udyog Vihar, Phase-IV, Gurugram-122015, Haryana, India  
Tel.:(+91) 124-454-0907, Toll Free No : (+91) 1800 10 23105 Website: www.sdbiosensor.co.in : Email: care@sdbiosensor.co.in  
CIN - U33110HR2012PTC047034

---

**Operations Qualification for Standard F 200**

1. Verified Cal 1 Calibration (blank check) : Done
2. Verified Cal 2 Calibration (UV light source and detector check) : Done
3. Verified Cal 3 Calibration (RBG light source and three level band Intensity (Strong positive, weak positive and negative) check) : Done

For **SD Biosensor Health Care Pvt. Ltd.**



Aamir Ayaz Khan

Assistant Product Manager



**SD BIOSENSOR HEALTHCARE PVT. LTD.**

Registered Office cum R&D Center : Plot No. 63, Udyog Vihar, Phase-IV, Gurugram-122015, Haryana, India  
Tel.:(+91) 124-454-0907, Toll Free No : (+91) 1800 10 23105 Website: www.sdbiosensor.co.in : Email: care@sdbiosensor.co.in  
CIN - U33110HR2012PTC047034

**Performance Certificate – Standard F 200**



**Instrument Certification:**

Study data has determined that the instrument described in this document either meets all criteria outlined in this Performance Qualification Protocol or exceptional conditions have been identified and documentation included.

It is recommended to do Instrument calibration by calibration set once in 30 days for checking optical parts of analyzer. There is no preventive maintenance required in this equipment. SD Biosensor will update software of system free of cost as when required to run machine at optimum levels.

This instrument is ready for specific usage.

Protocol Performed By:

Aamir Ayaz Khan

Assistant Product Manager



**SD BIOSENSOR HEALTHCARE PVT. LTD.**

Registered Office cum R&D Center : Plot No. 63, Udyog Vihar, Phase-IV, Gurugram-122015, Haryana, India  
Tel.:(+91) 124-454-0907, Toll Free No : (+91) 1800 10 23105 Website: www.sdbiosensor.co.in : Email: care@sdbiosensor.co.in  
CIN - U33110HR2012PTC047034

### In Field Calibration Report

Standard F200
Serial No : FA20FAITG1572




Operator
Name: S D Biosensor Healthcare Pvt Ltd

Instrument Owner
Name: Lupin Diagnostics, Jalna

Control Device	Lot No.	Expiry Device
CAL 1	67237M1F1BE02	20.07.2024
CAL 2	67237M1F1BE02	20.07.2024
CAL 3	67237M1F1BE02	20.07.2024
C -A1C - 1	9E129VA1C	09-05-2024
C - A1C -2	9E129V1AC	09-05-2024

Operating Condition	
Temperature	15-32 °C

Result of Control Device Measurements			
Control Device	Measured Value		Accepted Range
	1	2	
CAL 1	Ok	Ok	N/A
CAL 2	Ok	Ok	N/A
CAL 3	Ok	Ok	N/A
9E129V1AC - L1	5.8	5.9	4.3 % - 6.7 %
9E129V1AC - L2	9.3	9.4	7.5 % - 11.5 %

Final Result		
The Operator hereby confirm that the In Field Calibration was performed in accordance to the instructions given in " Standard F200 , In Field Calibration Procedure. This certificate is valid for 100 days.		
Date : 22.12.2022  Callibration Due Date : 30.03.2024	Authorized Operator Signature 	Instrument Responsible Signature  





# STANDARD™

2023-12-22

09:40

SN : FA20FAITG1461

Version : V001.064.16

Calibration Result

Operator : AJAY

Result : OK

Note :



# Detailed Result

[Calibration]

Date: 2023-12-22 09:40:32  
Operator ID: AJAY  
CAL-1 OK  
CAL-2 OK  
CAL-3 OK

OK

OK

Print

10 | 02:03:23 PM

SD BIOSENSOR

