

Calibration Certificate

Following voltages have been checked & adjusted on DiaSys SYS 200 Pro (Sl. No. SYS20220401) at Krsnaa Diagnostics – Rajpura, Punjab

Calibration Report:

System voltages are in the specified range – Checked by Digital Multimeter are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Power supply board o/p	+24.1 VDC	+24.0 +/- 0.5 VDC
		+05.0 VDC	+05.0 +/- 0.2 VDC
2	Lamp voltage	+12.1 VDC	12.0 +/- 0.5 VDC

System Temperature values are in the specified range –Checked by Digital Thermometer are as below:

S.No.	Module	Observed Value	Acceptable Range
1	Reaction Disk Temp	37.0 C	37.0 C +/- 0.3 C

Photometer Adjustment Values are within range as recommended:

S.No.	Module	Observed Value	Acceptable Range
1	ADC Count for all Wavelengths	10,240 to 10,300	8000 to 18,000

- All other mechanisms have been checked through field service program & Software, found OK.
- Performed calibration for End Point, Kinetic & Fixed Time Kinetic. Run control & sample. Results were satisfactory.
- Next Calibration will be performed on 28/05/2024.

For DiaSys Diagnostics India Pvt. Ltd.



Asmat Sami

Date: 28/05/2023

Regional Service Manager

www.DiaSys-Diagnostic.com

Contact: 022 3371 4323

Toll Free: 1800 120 1447

Registered Office

Plot No. A-821, T.T.C. Industrial Area,
MIDC, Mahape, Navi Mumbai - 400710
Maharashtra, India

DiaSys Diagnostics India Private Limited

The Specialists for diagnostic system solutions
Point-of care testing and automated Laboratory systems

CIN : U74999MH2013PTC248371

Email : info@diasys.in

Tel. No. : +91-022-71068331

Toll Free : 1800 120 1447

www.diasys-diagnostics.com