



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2023



PC-1024

Lab Name **MEDRAY CLINICS PVT LTD** Lab No **17475**
 Constituent Group **Chemistry I** Date of Result Entered : **20/11/2023**
 PT Item **Lyophilized human serum based** Date of Report Published : **05/12/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation)	1139	93.46	6.69	6.25	88 mg/dL	-0.87	0.37
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	711	91.14	8.85	8.06	109.4 mg/dL	2.26	0.60
3	CREATININE	Jaffes End point	Any Analyser (Automation / Semi Automation)	338	1.11	12.49	0.14	1.35 mg/dL	1.73	0.02
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	998	7.38	12.06	0.89	7.76 mg/dL	0.43	0.06
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	1156	4.13	10.84	0.45	4.03 g/dL	-0.22	0.03
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	823	2.55	9.64	0.25	2.57 g/dL	0.08	0.02
7	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	938	8.17	7.36	0.60	10.2 mg/dL	3.38	0.04
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	1000	3.67	15.36	0.56	4.02 mg/dL	0.62	0.04
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	1071	73.71	10.36	7.64	80.45 mg/dL	0.88	0.47
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	1070	166.27	7.19	11.95	166.49 mg/dL	0.02	0.73
11	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	836	20.98	13.04	2.74	23.1 mg/dL	0.77	0.19
12	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1140	111.95	13.93	15.59	128 U/L	1.03	0.92
13	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1129	29.54	15.14	4.47	38.2 U/L	1.94	0.27
14	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	936	105.34	11.90	12.53	152.2 U/L	3.74	0.82

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 MEDRAY CLINICS PVT LTD
 NO.962, 12TH MAIN ROAD, NEAR RELIANCE DIGITAL HALL 2ND STAGE, INDIRA NAGAR
 BANGALORE
 KARNATAKA560008

External Quality Assurance Scheme - Print Monthly Summary

Pamela Christudoss

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

28

Repeated

MEDRAY DIAGNOSTICS
INDIRANAGAR

No. : 16 Sample ID:EQAS REPEAT NOV Bed No. : Sample Type:Serum
Gender:Male Age: Case No. :
Doctor: Diagnosis:

Register Date:02-01-2024 17:04:08

No.	Test Name	Result	Hint	Unit	Reference range
1	[CAL1]Calcium	7.99	H	mg/dl	0.00-1.00
2	[ALP1]ALP	123.4		U/L	40.0-406.0

Checker:admin Examiner: BANGALORE Report Date:06-12-2023 17:28:00

A d d

Investigation checklist/Form

Survey information:				
Survey Name	EAAS			
Date survey received	Aug-23	Analyzer name/Model	AUTOCHEM XPERT	
Date survey result submitted	20-11-2023	Date of analysis performed	20-11-2023	
Investigation performed by		Date of report receipt	20-12-2023	
		Date of evaluation	05-12-2023	
Unacceptable parameter Name:				
Specimen	Analyte	Reported value	Repeated value	Date of retesting:
EAAS sample	Calcium	10.2	7.99	Intended/peer group value
				8.17

Root cause Analysis

Clerical	Yes	No	NA
1. Was the results correctly transcribed from instrument readout or report?	✓		
2. Was the correct instrument /method reagent reported on the result form?	✓		
3. Does the result reported on the result form match the result found on the proficiency testing evaluation report?	✓		
Procedural			
1. Was the written procedure followed?	✓		
2. Were the reagents within their open stability limit during analysis?	✓		
3. Were Quality Control results acceptable and without bias?	✓		
4. Were dilutions performed correctly?	✓		
Analytical			
1. Was the most recent calibration acceptable and within established limits at the time of testing?	✓		
2. Does a review of the past proficiency testing results indicate evenly distributed data without bias?	✓		
3. Was the intended result within measuring range for the instrument?	✓		
4. Was instrument maintenance performed on schedule?	✓		
5. Does a review of records indicate that there were no related instrument test problems noted prior to or after the proficiency testing as performed?	✓		
PT /EQAS material			
1. Was proficiency testing material received in the laboratory within an appropriate time after shipment?	✓		
2. Was proficiency testing material received at the appropriate temperature?	✓		
3. Were results graded in the appropriate peer group based on the method reported on the result form?	✓		

Conclusion /Summary:

Type of error		Survey evaluation problem
Method related		
Technical process related		✓
Clerical		

Preventive actions (If any) *After calibration again sample cal process.*

Review and approval:

A. [Signature]
6-12-2023

Investigation checklist/Form

Survey information:

Survey Name	EQAS	Analyzer name/Model	AVTO CHEM XPERT
Date survey received	Aug - 23	Date of analysis performed	20-11-2023
Date survey result submitted	20-11-2023	Date of report receipt	20-11-2023
Investigation performed by		Date of evaluation	05-12-2023
Unacceptable parameter Name:		Date of retesting:	
Specimen	Analyte	Reported value	Repeated value
EQAS SAMPLE	ALP	152.2	123.4
			Intended/peer group value
			105.3

Root cause Analysis

Clerical	Yes	No	NA
1. Was the results correctly transcribed from instrument readout or report?	✓		
2. Was the correct instrument /method reagent reported on the result form?	✓		
3. Does the result reported on the result form match the result found on the proficiency testing evaluation report ?	✓		
Procedural			
1. Was the written procedure followed?	✓		
2. Were the reagents within their open stability limit during analysis?	✓		
3. Were Quality Control results acceptable and without bias?	✓		
4. Were dilutions performed correctly?	✓		
Analytical			
1. Was the most recent calibration acceptable and within established limits at the time of testing?	✓		
2. Does a review of the past proficiency testing results indicate evenly distributed data without bias?	✓		
3. Was the intended result within measuring range for the instrument?	✓		
4. Was instrument maintenance performed on schedule?	✓		
5. Does a review of records indicate that there were no related instrument test problems noted prior to or after the proficiency testing as performed?	✓		
PT /EQAS material			
1. Was proficiency testing material received in the laboratory within an appropriate time after shipment?	✓		
2. Was proficiency testing material received at the appropriate temperature?	✓		
3. Were results graded in the appropriate peer group based on the method reported on the result form?	✓		

Conclusion /Summary:

Type of error

Method related		Survey evaluation problem	
Technical process related		Other (define below)	✓
Clerical			

Preventive actions (If any)
Review and approval:

After calibration again sample was repeated.

A ~~_____~~
6-12-2023