

**VIEW LAB MONTHLY SUMMARY**

<b>Lab Name</b>	<b>MEDRAY CLINICS PVT LTD</b>	<b>Details About Robust Analysis</b>
<b>Lab No</b>	<b>17475</b>	<b>Detail About Monthly Summary</b>
<b>Month</b>	<b>November</b>	<b>Detail about SDI</b>
<b>Year</b>	<b>2023</b>	
<b>Constituent Group</b>	<b>HbA1c</b>	

Click on the analyte to view Graphical Data

Date of Result Entered : 20/11/2023

Date of Report Published : 05/12/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	HbA1c	HPLC-Ion Exchange	Any Analyser	174	6.28	8.82	0.55	7.2 %	1.66	0.08

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>