

HIN/MED/2023-2024 /1055790  
20 TH JULY 2023

## **CALIBRATION CERTIFICATE**

This is to certify that the Hematology Analyzer Horiba YUMIZEN H550 bearing serial number 204YAXH03742 installed at **KRSNAA DIAGNOSTIC LAB GURDASPUR** as calibrated on 20<sup>TH</sup> JULY 2023.

Calibrator : ABX MINOCAL  
Lot No. : CX484  
Expiry Date : 05<sup>th</sup> AUG 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 19 JULY 2024



Shrish Dixit

(Head- Products & Marketing)  
For **Horiba India Pvt. Ltd.**

