



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7/3P/01



CC-2024

Certificate No	AACPL / 06363F	Field	Mech	Group	Speed	Page No.	1 of 1
SRF No & Date	2312216 DT - 21 / 12 / 2023			ULR No.	CC2646230000 0539F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krenaa Diagnostics Ltd			Calibration Date	21 / 12 / 2023		
Address	Civil Hospital Gurdaspur, Babri - 143520			Recommended Date for Next Calib	20 / 12 / 2024		
				Certificate Issue Dt.	23 / 12 / 2023		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / Sl. No	Party ID / Location
1	Centrifuge Machine	0 to 5000 RPM	100 RPM	Remi ZDIN-16612	KDDHPBCCF06 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/Sl No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / —	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS 12508 - 2018	RPM Measurement	56 %RH	25.7 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No. CP - 08

Calibration Result

Sl	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in RPM)
	UUC (in RPM)	(in RPM)	(in RPM)	
1	100	100.5	-0.5	57.75
2	500	500.9	-0.9	57.75
3	1000	1001.6	-1.6	57.75
4	2500	2502.7	-2.7	57.77
5	5000	5004.6	-4.6	57.77

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) supplied for calibration
- UUC - Unit under calibration

— End of Certificate —

Calibrated by
[Signature]

Approved by
[Signature]
Shoresh Kumar
(Tech. Manager)