



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7/21/2023



CC-2546

Certificate No	AACPL / 09324F	Field	Mech	Group	Volume	Page No.	1 of 1
SRF No & Date	2312214 DT - 21 / 12 / 2023			ULR No.	CC2646230000 09324F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krishna Diagnostics Ltd			Calibration Date	21.12.2023		
Address	Civil Hospital Gurdaspur Babri - 143520			Recommended Date for Next Callb.	20.12.2024		
				Certificate Issue Dt.	23.12.2023		
				UUC Condition	Good		

Calibration Instrument Detail

Sr	Instrument Name	Range	Least Count	Make / Sl.No.	Party ID / Location
1	Micro Pipette	20 to 200 µl	1 µl	Dragon Lab YE222AX0047997	KDDHPBCCMF07 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

Sr	Instrument Name	Make/Sl No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668663	Bluebox Techno	BBT452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 6655 - B	Volume Measurement	54 %RH	20.4 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C Section No. CP - 03

Calibration Result

Sr No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in µl)
1	20	19.99856	0.00144	0.25
2	100	99.99514	0.00486	0.25
3	200	199.99121	0.00879	0.25

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - unit under calibration

----- End of Certificate -----

Asht
Calibrated by



Approved by
Arind Chaurasia
Arind Chaurasia
(CEO)