



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.07.01



CC - 2548

Certificate No	AACPL / 09325F	Field	Mech	Group	Volume	Page No.	1 of 1
SRF No & Date	2312214 DT - 21 / 12 / 2023		ULR No.		CC2548230000 09325F		
Calibration Method	Comparison Method		Accuracy				
Company Name	M/s Krishna Diagnostics Ltd		Calibration Date		21.12.2023		
Address	Civil Hospital Gurdaspur, Babri - 143520		Recommended Date for Next Calib		20.12.2024		
			Certificate Issue Dt.		23.12.2023		
			UUC Condition		Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / Sl.No.	Party ID / Location
1	Micro Pipette	100 to 1000 μ l	5 μ l	Dragon Lab	KDDHPBCCMP08 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighting Balance	Radweg / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO 8555 - 6	Volume Measurement	55 %RH	20.3 C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No. CP / 03

Calibration Result

S. No.	Nominal Value on UUC (in μ l)	Average Volume at 27 °C (in μ l)	Deviation (in μ l)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in μ l)
1	100	100.00582	-0.00582	0.25
2	500	500.01152	-0.01152	0.25
3	1000	1000.01689	-0.01689	0.25

Note

- The calibration results reported in this calibration certificate are valid at the time of issue under stated condition.
- This certificate cannot be reproduced except in full without our prior permission in writing.
- This certificate refers only to the particular item(s) submitted for calibration.
- UUC - Unit under calibration.

----- End of Certificate -----

Asish
Calibrated by



Approved by
Asish
Asish Chaurasia
(CEO)