



# A A CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7 34-01

CC - 2645

Certificate No	AACPL / 09326F   Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	2312214 DT - 21 / 12 / 2023		ULR No.	CC2646230000 09326F		
Calibration Method	Comparison Method		Accuracy			
Company Name	M/s Krishna Diagnostics Ltd.		Calibration Date	21 12 2023		
Address	Civil Hospital, Gurdaspur, Babri - 143520		Recommended Date for Next Calib	20 12 2024		
			Certificate Issue Dt.	23 12 2023		
			UUC Condition	Good		

### Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	5 to 50 µl	0.5 µl	I-Pipette	KDDHPBCCMP09 Serology

### Standard Equipments Used ( Traceable to National Standard )

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radway / 568583	Bluebox Techno	BBT/452/JAN/23	05-03-2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8555 - 5	Volume Measurement	55 %RH	20.5 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No CP - D3

### Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Deviation (in µl)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in µl)
1	10	10.00255	-0.00256	0.25
2	25	25.00714	-0.00714	0.25
3	50	50.01132	-0.01132	0.25

#### Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition.
- This certificate cannot be reproduced except in full without our prior permission in writing.
- This certificate refers only to the particular items submitted for calibration.
- UUC - Unit Under Calibration.

— End of Certificate —

*Asht*  
Calibrated by

Approved by  
*Asht Chaurasia*  
Asht Chaurasia  
(CEO)