



The Accurate Calibration & Medical Solution

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF MICROSCOPE

Certificate No :	TAC/12-23/KDL/MS-1001	Calibration Date :	21.12.2023
Job Identification No :	TAC/C/12/12/2023	Next Calibration Date :	20.12.2024
Ref. No :	SRF Dated - 18. 12. 2023	Issue Date :	23.12.2023
Customer Name Address	M/S KRSNAA DIAGNOSTICS LTD. CIVIL HOSPITAL GURDASHPUR, BABRI-143520		

DESCRIPTION OF THE EQUIPMENT TO BE CALIBRATED

Instrument Name	MICRO SCOPE		
Make	GLOMEX	Visual Inspection	OK
Model	-----	I.D. No.	KDDHPBCPM505
Serial No.	-----	Location / Dept.	HAEMATOLOGY

CALIBRATION STANDARD USED

Sl.No.	Master Equipment	Manufacturer	Model No.	Serial No.	Date of Calibration	Valid Upto
1	Electrical Safety Analyzer	DATREND Systems Inc.	vPed - Mini	VPM18100028	14.10.2023	14.10.2024

ENVIRONMENTAL CONDITIONS	Temperature	22.0 °C ± 2.0 °C	Humidity	50 ± 10%RH
--------------------------	-------------	------------------	----------	------------

RESULT

VISUAL TEST

Sr.No.	Tests	Remarks
1	Power Cords/Cable Checking	OK
2	Mains Socket Checking	OK
3	Equipment Type (B,BF,CF)	CF
4	Equipment Class (I, II, III)	I

FUNCTIONAL TEST

Sr.No.	Parameters	Measured Value	Remarks
1	Voltage Between Live and Neutral (VLN)	231.2 V (220-240 V)	OK
2	Voltage Between Live and Earth (VLE)	230.2 V (220-240V)	OK
3	Voltage Between Neutral and Earth (VNE)	0.4 V (0-5V)	OK
4	Load Current (Ia)	0.6 A	OK
5	Enclosure Current	0.7 µA (0-10µA)	OK
6	Leaking Current (IL)	112 µA (0-300µA)	OK
7	Time for Testing (t)	15 (min)	OK

Electrical Safety Test Passed as per IEC Standard 60601-1

Calibration Passed

YES



NO



Remarks:

1. Certificates shall not be reproduced except the written approval of The Accurate Calibration & Medical Solution.
2. The Calibration result reported in this certificate are only valid under the conditions of calibration and the stated condition of measurement.

Calibrated by



Approved by

Plot No. 72/G1, Ganesh Vihar Sirsi Road, Jaipur (Raj.) -302012
Mob.: 9672294758

E-mail : theaccuratecalibration@gmail.com