



# R-TECH CALIBRATION

ISO/IEC 17025:2017 - NABL ACCREDITED LABORATORY

52/1, Pathak Para Road, Behala, Room No. G1, Kolkata - 700 060



CC-3301



## CALIBRATION CERTIFICATE

ULR :	CC330124000000008F		CERTIFICATE NO.:	RTC/24/008	
Date of Receipt	Date of Calibration	Certificate Issue Date	Location of Calibration	Suggested Due Date	SRF No./ Date
03-Jan-24	03-Jan-24	05-Jan-24	SITE	02-Jan-25	SRF/23/002

Name of Customer : M/s	GOODDAYS DIAGNOSTIC CENTRE		Environmental Conditions		
	Farmside Road, Kodalia Gram Panchayat, Chinsurah, Hooghly Pin-712 102, West Bengal		Temperature: 25 ± 4°C	Relative Humidity: 50 ± 20% RH	

Description of DUC	Centrifuge		Make	Remi	Range	3500 RPM	
Name of Equipment:	Centrifuge		Modle	C-854/6	L.C.	1 RPM	
Condition of Item:	Good	Calibration Discipline :	Mechanical	ID. No.	Type	Analog	
Location of DUC :	N/M	Calibration Parameter :	Speed	Sr. No.	ZEDN-14978	Accuracy	N/M

### REFERENCE STANDARD & MEASUREMENT TRACEABILITY

Name	Serial No.	Certificate No.	Calibration Date	Due Date	Traceability No.
Dig. Tachometer	1610208981	ATL/E/220523/009	15-May-23	15-May-24	CC-2590

Calibration Procedure No.: The Calibration is carried out as per the "RTC-CP-M-04".

### CALIBRATION RESULT

Sr. No.	Calibration Point	Avg. Standard Reading	Avg. DUC Reading	Deviation	Expanded Uncertainty
	(RPM)	(RPM)	(RPM)	(RPM)	(RPM)
1	500	500.6	500	-0.6	±16
2	1000	1002	1000	-2	±16
3	2000	2003	2000	-3	±16
4	3000	3001	3000	-1	±16
5	3500	3503	3500	-3	±16

Note: \*DUC: Device Under Calibration , N/M : Not Mentioned.

The Error is positive except when indicated otherwise. It has to algebraically subtracted from DUC Reading in order to get the standard reading.

The Expanded Uncertainty associated with measurement at 95.4% confidence level & coverage factor K = 2 is above under uncertainty column.

Note :-

1. This certificate refers only to the particular item(s) submitted for calibration at Site/Laboratory.
2. Results reported are valid at the time of and under the stated conditions of measurement.
3. Certificate should not be reproduced, except in full, without the prior permission of Laboratory.
4. Next Calibration due date mentioned as per customer requirement.
5. All reference standard used for calibration are traceable to National Standard.

\*\*\*\*\* End of Certificate \*\*\*\*\*

*Chandan*



*Ramy*

Calibration Engineer  
(Chandan)

Authorized Signatory  
(Ramasare Prajapati)

Format : RTC-QF-7.8-02

Page No.: 01/01



033-3557 4693



7044226000 / 7891753269



rtechgroups06@gmail.com  
lab@rtechcalibration.in  
info@rtechcalibration.in



http://rtechcalibration.in

CALIBRATION

SALES & SERVICE

REPAIR