



HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **MICROS ES60** bearing serial number: **110ESOH18819** installed at **KRSNAA DIAGNOSTICS, RAJPURA, PUNJAB** has calibrated on 13TH FEB 2023.

Calibrator : ABX MINOCAL
Lot No. : CX479
Expiry Date: 05/03/2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on **12TH FEB 2024**.



Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**