

A A CALIBRATION PVT. LTD. CALIBRATION CERTIFICATE



CC - 264

Certificate No	AACPL / 07156F Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	230930.2 DT - 30 / 09 / 20	23	ULR No.	CC26	46230000 0715	
Calibration Method :	Comparison Method		Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd. SDH Rajpura, A.P Jain Civil Hospital,		Calibratio	30.09	9.2023	
Address			Recommend	ext Calib. 29.09	9.2024	
	Rajpura (PB)		Certificate	Issue Dt.	03.10	0.2023
			UUC Cond	dition	Good	1

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI. No	Party ID / Location
1	Micro Pipette	20 to 200 μl	1 µl	I-Pette	KDSDPBCCMP10
					Bio Chemistry

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/452/JAN/23	05/03/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	(50 ± 10)%RH	(23 ± 2)°C	Laboratory

Calibration Procedure: As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

	Calibration Result					
S.	Nominal Value on UUC	Value on UUC		Uncertainty at approx 95% confidence		
No	(in µl)	(in µl)	(in µI)	Level and coverage factor k = 2 is ± (in μl)		
1	20	20.00274	-0.00274	0.25		
2	100	100.00563	-0.00563	0.25		
3	200	200.00689	-0.00689	0.25		

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC Unit under calibration

----- End of Certificate -----

Asluck Calibrated by CALIBRATED 5

Approved by

Anada Chaurasia (CEO)