



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2023



PC-1024

Lab Name	<b>TIRUVALLA MEDICAL MISSION HOSPITAL</b>	Lab No	<b>0053</b>
Constituent Group	<b>HbA1c</b>	Date of Result Entered :	<b>11/12/2023</b>
PT item	<b>Lyophilized human whole blood based</b>	Date of Report Published :	<b>02/01/2024</b>

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	Turbidimetric Inhibition Immunoassay	Roche C111, C311, Integra 400 plus, C501, C502, C702	142	5.27	6.35	0.34	<b>5.67 %</b>	1.19	0.06

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

**TIRUVALLA MEDICAL MISSION HOSPITAL  
TIRUVALLA  
TIRUVALLA  
KERALA689101**

**Coordinator Contact Details:**  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
**Dr. Pamela Christudoss**  
**CMC EQAS Coordinator**  
**Christian Medical College, Vellore**

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*