

CALIBRATION CERTIFICATE

This is to Certify that the Mechanical, Electronics and Software Calibration has been done on 28 April 2023 for SYS-200 Pro FullyAutomatic Analyzer having serial number (SYS20220204) carried out in Krsnaa Diagnostics – Batala, Punjab

**Following Voltage have been checked & adjusted on SYS 200 Pro:-
DC Power Supply:**

Target Volt	Range	Actual Volt
+12 V	+/-0.5 V	12.06 V
+5 V	+/- 0.2 V	5.02 V
+24 V	+/- 1.0 V	24.10 V

- **Checked the temperature of Reagent Preheater, Reaction Disk & Reagent disk Cooling.**
- **All other mechanism have been checked through software and found OK.**
- **Light Quantity Check up and Cuvette Blank Tests for all the wavelengths are checked and found with in the Range.**
- **Performed calibration for End Point, Kinetics & Fixed Point Kinetics parameters (Glucose, Creat, ALT, etc).**
- **Run Controls and results were Satisfactory.**
- **Next Calibration will be performed on 28/04/ 2024**

For Diasys Diagnostics India Pvt Ltd .



**Sami Asmat
Reginal Service Manager**

Date :28 April 2023