



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



Lab Name DELHI DIAGNOSTICS
Constituent Group Chemistry II
PT item Lyophilized human serum based

Lab No 16240
Date of Result Entered : 15/07/2023
Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1579	99.40	10.09	10.03	109.1 mg/dL	0.97	0.50
2	UREA II	Others (Urease Berthelot / Ned Dyc) II	Any Analyser (Automation / Semi Automation)	760	139.56	10.79	15.06	153.4 mg/dL	0.92	1.09
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1221	2.38	12.42	0.30	2.15 mg/dL	-0.78	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1399	3.16	18.49	0.58	2.92 mg/dL	-0.41	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1311	6.03	10.34	0.62	6.3 g/dL	0.43	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1273	3.57	10.15	0.36	3.6 g/dL	0.08	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1443	6.23	17.55	1.09	7.8 mg/dL	1.44	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1055	126.41	12.15	15.36	130.3 mg/dL	0.23	0.95
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1376	121.86	15.59	18.99	136.5 mg/dL	0.77	1.02

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
DELHI DIAGNOSTICS
1085, FIRST FLOOR, PURUN ENCLAVE, 29-29 CHOWK, OLD FARIDABAD
FARIDABAD
HARYANA121002

*All result within range
No need CAPA
N.K. Shyam*

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name DELHI DIAGNOSTICS Lab No 16240
 Constituent Group Chemistry II Date of Result Entered : 14/08/2023
 PT item Lyophilized human serum based Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1614	179.75	10.04	18.05	181.4 mg/dL	0.09	0.90
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	805	24.94	14.69	3.66	26.5 mg/dL	0.43	0.26
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1202	6.39	12.07	0.77	6 mg/dL	-0.51	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1330	2.21	19.85	0.44	2 mg/dL	-0.48	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1298	5.71	10.24	0.58	6.1 g/dL	0.67	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1236	3.41	9.09	0.31	3.3 g/dL	-0.35	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1383	6.64	13.75	0.91	6.4 mg/dL	-0.26	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1501	121.44	9.98	12.12	120 mg/dL	-0.12	0.63
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1321	134.70	12.22	16.46	136 mg/dL	0.08	0.91

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 DELHI DIAGNOSTICS
 1085, FIRST FLOOR, PURUN ENCLAVE, 29-29 CHOWK, OLD FARIDABAD
 FARIDABAD
 HARYANA 121002

*All result within range
 No Need CAPA
 N.K. Sharma*

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

10/9/23, 12:59 PM



PC-1024

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023

Lab Name **DELHI DIAGNOSTICS**
 Constituent Group **Chemistry II**
 PT item **Lyophilized human serum based**

Lab No **16240**
 Date of Result Entered : **12/09/2023**
 Date of Report Published : **05/10/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1697	333.34	11.48	38.25	316 mg/dL	-0.45	1.86
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	913	38.17	11.97	4.57	38 mg/dL	-0.04	0.30
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1241	1.16	14.35	0.17	1.12 mg/dL	-0.24	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1395	5.18	15.03	0.78	4.8 mg/dL	-0.49	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1299	5.64	10.51	0.59	5.4 g/dL	-0.40	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1246	3.31	9.55	0.32	3.1 g/dL	-0.66	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1315	9.82	13.19	1.30	10.3 mg/dL	0.37	0.07
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1501	119.45	10.95	13.08	126 mg/dL	0.50	0.67
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1333	201.48	11.49	23.15	175 mg/dL	-1.14	1.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 DELHI DIAGNOSTICS
 1085, FIRST FLOOR, PURUN ENCLAVE, 29-29 CHOWK, OLD FARIDABAD
 FARIDABAD
 HARYANA121002

*nil report within range
 No need (APA)*

NK.Sharma

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - OCTOBER 2023



Lab Name DELHI DIAGNOSTICS
Constituent Group Chemistry II
PT item Lyophilized human serum based

Lab No 16240
Date of Result Entered : 14/10/2023
Date of Report Published : 04/11/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1673	129.03	9.72	12.54	130 mg/dL	0.08	0.61
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	920	84.21	11.70	9.85	76.5 mg/dL	-0.78	0.65
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1204	7.35	12.30	0.90	7.1 mg/dL	-0.28	0.05
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1325	8.35	11.84	0.99	8 mg/dL	-0.35	0.05
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1282	5.47	10.35	0.57	4.8 g/dL	-1.18	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1253	3.29	9.69	0.32	3.2 g/dL	-0.28	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1268	5.07	16.36	0.83	7.4 mg/dL	2.81	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1485	109.78	11.30	12.40	125.3 mg/dL	1.25	0.64
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1243	104.67	16.11	16.86	114 mg/dL	0.55	0.96

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
DELHI DIAGNOSTICS
1085, FIRST FLOOR, PURUN ENCLAVE, 29-29 CHOWK, OLD FARIDABAD
FARIDABAD
HARYANA121002

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

*All result within range
No need CAPA
N.K.Sharma*



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2023



Lab Name DELHI DIAGNOSTICS

Lab No 16240

Constituent Group Chemistry II

Date of Result Entered : 12/11/2023

PT item Lyophilized human serum based

Date of Report Published : 05/12/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1594	238.88	9.59	22.91	256 mg/dL	0.75	1.15
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	917	66.61	10.47	6.98	56.8 mg/dL	-1.41	0.46
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1210	4.52	12.57	0.57	5.1 mg/dL	1.02	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1211	1.15	23.45	0.27	1.1 mg/dL	-0.19	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1305	4.77	13.33	0.64	4.5 g/dL	-0.42	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1267	2.90	11.42	0.33	2.8 g/dL	-0.30	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1323	8.13	12.01	0.98	8.6 mg/dL	0.48	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1480	98.77	11.12	10.98	98.8 mg/dL	0.00	0.57
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1379	88.73	18.66	16.56	89.1 mg/dL	0.02	0.89

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

All result within range
No need CAPA
NK Sharma

LAB ADDRESS :
DELHI DIAGNOSTICS
1085, FIRST FLOOR, PURUN ENCLAVE, 29-29 CHOWK, OLD FARIDABAD
FARIDABAD
HARYANA121002

Coordinator Contact Details:
Email:clinq@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2023



PC-1024

Lab Name **DELHI DIAGNOSTICS** Lab No **16240**

Constituent Group **Chemistry II** Date of Result Entered : **17/12/2023**

PT Item **Lyophilized human serum based** Date of Report Published : **02/10/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1639	102.89	9.66	9.94	106 mg/dL	0.31	0.49
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	849	105.59	13.68	14.44	107 mg/dL	0.10	0.99
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	1228	1.79	12.56	0.22	1.9 mg/dL	0.49	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Lendrassik II	Any Analyser (Automation / Semi Automation)	1302	4.67	13.51	0.63	4.25 mg/dL	-0.67	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1278	5.22	10.56	0.55	5 g/dL	-0.40	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1248	3.16	10.10	0.32	2.9 g/dL	-0.82	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1382	5.59	15.59	0.87	5.6 mg/dL	0.01	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1470	106.78	10.69	11.42	105.2 mg/dL	-0.14	0.60
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1327	144.29	12.48	18.01	150 mg/dL	0.32	0.99
SDI Range				Interpretation						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution, Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

all result within range

No need CAPA

N.K. Sharma

Pamela Christodoss

LAB ADDRESS :
DELHI DIAGNOSTICS
1085, FIRST FLOOR, PURUN ENCLAVE, 29-29 CHOWK, OLD FARIDABAD
FARIDABAD
HARYANA121002

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data In CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****