

Calibration report

Product Name	Semi auto Biochemistry Analyzer	Model	URA
Tested Date	25.02.2023	SN	601631019 E
Customer Name	DELHI DIAGNOSTICS		00101017

Parameter	expected result	Object to the	T
Flowcel Temp	37° ±0.3° Celsius	Obtained result	Conclusion
Flowcel aspiration	450ul ± 10%(450ul used as standard to	37.01° Celsius	₩/ NOT OK
volume	measure in each parameter program)	460ul	9₩/ NOT OK
340nm	AD values 2000 to 3800	2920	OK/ NOT OK
405nm	AD values 2000 to 3800	- 3020	OK/ NOT OK
500nm	AD values 2000 to 3800	3250	OK/ NOT OK
546nm	AD values 2000 to 3800	3050	ØR/ NOT OK
578nm .	AD values 2000 to 3800	3230	OK/ NOT OK
620nm	AD values 2000 to 3800	3090	OK/ NOT OK

esting Result:		DERING	The state of	en salara da	d		
Crity		S. F. Sale	00 5400				
emarks:		37 () 1 1 1 4 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	QC PASS	, to 4"			
strument Calibrated	as per facto	orv standa	ord and the nex	fa			
	יים איים ביים	ory scallud	ii u anu the ber	Tormed c	hack valu	100	
				·······································	HECK VAIL	ies are sat	tistactory
	,				TIECK Valu	sare sat	tisfactory
	,			-		es are sat	tisfactory
	,			-	THECK VAIL	res are sat	tisfactory
Next calibration on: 25	,		<u> </u>	-		uditerre sar	

Medsource Ozone Biomedicals Pvt. Ltd.

Head Office: Plot No. 109, HSIIDC, Sector-31, Faridabad-121003, (Haryana), INDIA Registered Office: 238, Sant Nagar, East of Kailash, New Delhi-110065, INDIA Telephone: +91-129-4286600 | E-mail: mail@ozonebio.com | Website: www.ozonebio.com CIN: U85100DL2003PTC120213



IQ/OQ/PQ Protocol

Installation Qualification

Operational Qualification

Performance Qualification

Instrument name: URA

Serial No. 601631019 IE

D.O.I.: 06th April 2015

Customer Details

Delhi Diagnostic

Puran Enclave, 28/29 Chowk 1st Floor , Near Pooja Furniture ,Faridabad Haryana 121002

Contact number - 8010639426

Email id- delhidiagnostics3456@gmail.com



Corporate office: Plot No109, HSIIDC, Sector-31, Faridabad - 121003 (Haryana), Phone: +91-129-4286600/612 Email mail@ozonebio.com, Website: www.ozonebio.com.



Installation Qualification

To verify that the facilities provided for the installation of semi automated Biochemistry analyzer meets the requirements for the system and that all components of the system have been delivered, ready for installation in accordance with the standard for the system and other appropriate regulations.

Purpose: The purpose of the installation qualification (IQ) is to confirm that the instrument is installed as per the requirement.

S.No.	Parameter	Expected Result	Observed	Remarks		
1	Ground voltage(voltage between neutral and earth)	< 3V	1.3 V	ок		
2	Line voltage (voltage between line and neutral)	180V to 245V	228 V	ок		
3	Capacity of UPS used	>400W	600VA	ок		
4	Installation of peristaltic tube	Installed	ОК	ок		
	Unpacking					
		No damage.	NO DAMAGE SEEN	CHECKED		
5	Package inspection	Inspection of accessory supplied		CHECKED		

- The analyzer passed the installation qualification
- The analyzer didn't pass the installation qualification

Special Note: Installation report attached.

Engineer name and signature

Corporate office: Plot No109, HSIIDC, Sector-31, Faridabad - 121003 (Haryana), Phone: +91-129-4286600/612 Email mail@ozonebio.com, Website: www.ozonebio.com.

Customer name and signature

Operational Qualification

Purpose: The purpose of the operational qualification (OQ) is to ensure that the system can be operated as expected and to devise the standard operating procedure.

S no	Parameter	Expected result	Result
1	Switch ON the instrument	Software started without any error message	NO ERROR SEEN
2	Functionality of peristaltic system	Di water aspirated without any bubble in fluidic path	CHECKED
3	Functionality of halogen lamp	Lamp is glowing	CHECKED
4	Functionality of touch panel	Able to operate the software with touch.	N.A.
5	Functionality of Internal printer	Responding to print command	CHECKED
6	Functionality of flowcel temperature	Temperature reach 37±0.4°C and stable	CHECKED
7	Programming of all reagents.	Installed	CHECKED
		Have all users has been properly trained on the safety, theory of operation and	
8	Training	maintenance of the URA?	ОК

- The analyzer passed the operational qualification
- O The analyzer didn't pass the operational qualification

Special note:

Engineer name and signature

Customer name and signature





Corporate office: Plot No109, HSIIDC, Sector-31, Faridabad - 121003 (Haryana), Phone: +91-129-4286600/612 Email mail@ozonebio.com, Website: www.ozonebio.com.