



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2023



PC-1024

Lab Name **HLM JALANA - LUPIN DIAGNOSTICS** Lab No **18900**
Constituent Group **HbA1c** Date of Result Entered : **14/11/2023**
PT item **Lyophilized human whole blood based** Date of Report Published : **05/12/2023**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	HPLC-Ion Exchange	Tosoh G8 - HPLC / HPLC-723G8)	210	6.37	3.89	0.25	5.9 %	-1.90	0.03

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

HLM JALANA - LUPIN DIAGNOSTICS
JALNA CRITICAL HOSPITAL, MAHALAXMI NAGAR, MANTHA CHWOK
JALNA
MAHARASHTRA431203

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****