



## Hemoglobin (Hb1Ac)

### ALL METHOD REPORT

Lab Code: 3041

Cycle - 12/2023  
Round No - 4

Date: 29/12/2023


Parameters	Units	No.of. Participants	Robust Mean	Robust Standard deviation (SD)	Uncertainty of Assign values	Range ( $\pm 2$ SD)	Your Value	Z Score
Hemoglobin (Hb A1C)	%	103	5.9	0.5	0.06	4.8-7.0	6.8	1.8

### Interpretation of Z Score :

Z Score Value(+/-)	$[Z] \leq 2.0$	$2.0 < [Z] < 3.0$	$[Z] \geq 3.0$
Interpretation	Satisfactory Performance No signal	Questionable Warning Signal	Unsatisfactory Performance action Signal

Legends	(*) Excluded From Group Mean	{.} Not Reported	(#)Late Result Submission	(\$ )Reported in other Unit
---------	------------------------------	------------------	---------------------------	-----------------------------


Chief Coordinator

  
Dr.Sanjay Mehrotra

Checked By:

Prepared by: NT

Programme Director

  
Dr.Bandana Mehrotra

Page 1 of 5

Doc. No.: ASS / FR / 11 / R 01 / Dt.: 10.01.2022





## Hemoglobin ( Hb A1C)

### METHOD WISE REPORT

Lab Code: 3041

Cycle - 12/2023  
Round No - 4

Date: 29/12/2023

Note: Your lab is not the part of method group

Parameters	Units	Method Group	No.of Participants	Robust Mean	Robust Standard deviation (SD)	Uncertainty of Assign Values	Range ( $\pm 2$ SD)	Your Value	Z Score
Hemoglobin (Hb A1C)	%	HPLC	46	5.7	0.3	0.06	5.0-6.4	-	-

#### Interpretation of Z Score :

Z Score Value(+/-)	$[Z] \leq 2.0$	$2.0 < [Z] < 3.0$	$[Z] \geq 3.0$
Interpretation	Satisfactory Performance No signal	Questionable Warning Signal	Unsatisfactory Performance action Signal

Legends	(*) Excluded From Group Mean	{.} Not Reported	(#)Late Result Submission	(\$)Reported in other Unit
---------	------------------------------	------------------	---------------------------	----------------------------

Chief Coordinator

Dr.Sanjay Mehrotra

Checked By:

Prepared by: NT

Doc. No.: ASS / FR / 11A / R 01 / Dt.: 10.01.2023

Programme Director

Dr.Bandana Mehrotra

Page 2 of 5



## Hemoglobin ( Hb A1C)

### METHOD WISE REPORT - PART-02

Lab Code: 3041

Cycle - 12/2023  
Round No - 4

Date: 29/12/2023

Note: Your lab is not the part of method group

Parameters	Units	Method Group	No. of Participants	Robust Mean	Robust Standard deviation (SD)	Uncertainty of Assign Values	Range ( $\pm 2$ SD)	Your Value	Z Score
Hemoglobin (Hb A1C)	%	Immunturbidimetric	22	6.2	0.9	0.24	4.5-7.9	-	-


#### Interpretation of Z Score :

Z Score Value(+/-)	$[Z] \leq 2.0$	$2.0 < [Z] < 3.0$	$[Z] \geq 3.0$
Interpretation	Satisfactory Performance No signal	Questionable Warning Signal	Unsatisfactory Performance action Signal

Legends	(*) Excluded From Group Mean	{ } Not Reported	(#) Late Result Submission	(\$) Reported in other Unit
---------	------------------------------	------------------	----------------------------	-----------------------------

Chief Coordinator

  
Dr. Sanjay Mehrotra

Checked By 

Programme Director

  
Dr. Bandana Mehrotra

Page 3 of 5







## Hemoglobin (HbA1C)

### GRAPHICAL REPORT

Cycle - 12/2023  
Round-4

Date: 29/12/2023

