

CERTIFICATE OF CALIBRATION

Customer Name & Address	District Women Hospital Jhansi Uttar Pradesh 284001 India	Certificate No.	Z3JDWH6
		Date of Calibration	28.11.2023
		Valid Upto	27.11.2024
		Date of Issue	29.11.2023

Instruments Details			
Instrument Name	Refrigerator	S No.	na
Make/Model No.	Videocon/s169	Location	PATHOLOGY
ID No.	MHUP - 401-0496	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	THERMO HYGROMETER	ELECTRICAL SAFETY ANALYZER	DIGITAL THERMOMETER
Make	HTC	FLUKE	METRAVI
Model/SR No.	288-ATH/NA	ESA612/2959012	DTM-900/200923407
Calibration Validity	20/07/2024	23/01/2024	20/07/2024
Calibration By	UNIQUE CALIBRATION SOLUTIONS	STIC	UNIQUE CALIBRATION SOLUTIONS

Environmental Condition			
Temperature	27 °C	Relative Humidity	56 %

Calibration Results				
SL NO	SET READING (°C)	MEASURED READING(°C)	TOLERANCE(°C)	STATUS
1	4	3.2	±2	P
2	4	4.5	±2	P
3	4	4.6	±2	P

ELECTRICAL SAFETY REPORT ATTACHED

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

KHAN AMIR



Approved By

ASWANTH TM

