



Quality Services & Laboratories

Plot No.10, Second Floor, D.S.I.D.C. Scheme - III
Okhla Industrial Area, Phase - II, New Delhi- 110020, India,
Tel. : +91 85953 43496
Email: corp@qsl.co.in ; Web.: www.qslglobal.com



CC-2900

CALIBRATION CERTIFICATE

Unique Lab Report No. : ULR-CC290023000009532F Page 1 of 1
Certificate No. : QSL/DEL/CAL/23122901.05 SRF No. : QSL/23122901
Date of Calibration: 29.12.2023 Suggested Due Date: 29.12.2024
Date of Receipt of Instruments: 27.12.2023 Date of Issue: 30.12.2023

Customer Name & Address:

M/s: National Clinical Laboratory,
Opp. R.K Joshi Government Hospital,
Dausa, Distt. DAUSA Rajasthan - 303303.

Description of Instrument:

Name :	Micro Pipette	Range :	10-100 μ l
Make :	Bexco	Least Count :	---
S. No. :	MVV100	Performed at :	Lab
Model No. :	---	Visual Inspection :	OK
Location:	Microbiology and Infectious Disease Serology	Zero Error :	Nil
Customer ID:	NCL/MS/MP/02		

Environmental Condition:

Temperature :	25°C \pm 3 °C	Humidity :	40% to 60%
---------------	-----------------	------------	------------

Methodology of Calibration:

Reference Standard: IS: 4787	Work Instruction No. :	QSL/WI/M/21
------------------------------	------------------------	-------------

Standard & Major Equipment (s) Used for calibration:

S.No.	Instrument Name	Make	S. No./Model No.	Certificate No.	Calibrated By	Valid Upto
1.	Std. Weight Box	Weightronics	---	WMCL/F/2023-06/2469	Weightronics	08.06.2025

Calibration Result:

UUC Reading in μ l	Standard Reading in μ l	Error in μ l	Uncertainty of Measurement:
10	10.123	-0.123	\pm 0.9 μ l
20	20.156	-0.156	\pm 0.9 μ l
50	50.252	-0.252	\pm 0.9 μ l
80	80.269	-0.269	\pm 0.9 μ l
100	100.364	-0.364	\pm 0.9 μ l

(The reported expanded Uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor $k = 2$, which for a normal distribution corresponds to a coverage probability of approximately 95%)

Note:

- *The calibration results reported in this calibration certificate are valid at the time of & under stated conditions of measurement.
- *This certificate cannot be reproduced except in full without our prior permission in writing.
- *This certificate refers only to the particular item (s) calibrated.
- *Laboratory standard are traceable to national standards.
- *UUC: Unit Under Calibration

Prepared By

Customer Service Cell (Gajendra Singh)

Approved By

CEO (Mdul Kohli)

QSL-2G-315

All business are undertaken subject to company standard general business conditions, which are available on request.