



# Quality Services & Laboratories

Plot No.10, Second Floor, D.S.I.D.C. Scheme - III  
Okhla Industrial Area, Phase - II, New Delhi- 110020, India,  
Tel. : +91 85953 43496  
Email: corp@qsl.co.in ; Web.: www.qslglobal.com



CC-2900

## CALIBRATION CERTIFICATE

Unique Lab Report No. : ULR-CC290023000009531F Page 1 of 1  
Certificate No. : QSL/DEL/CAL/23122901.04 SRF No. : QSL/23122901  
Date of Calibration: 29.12.2023 Suggested Due Date: 29.12.2024  
Date of Receipt of Instruments: 27.12.2023 Date of Issue: 30.12.2023

### Customer Name & Address:

M/s: National Clinical Laboratory,  
Opp. R.K Joshi Government Hospital,  
Dausa, Distt. DAUSA Rajasthan - 303303.

### Description of Instrument:

Name :	Micro Pipette	Range :	100-1000 µl
Make :	Bexco	Least Count :	-----
S. No. :	----	Performed at :	Lab
Model No.:	MVVV1000	Visual Inspection :	OK
Location:	Microbiology and Infectious Disease Serology	Zero Error :	Nil
Customer ID:	NCL/MS/MP/03		

### Environmental Condition:

Temperature :	25°C ± 3 °C	Humidity :	40% to 60%
---------------	-------------	------------	------------

### Methodology of Calibration:

Reference Standard: IS: 4787	Work Instruction No. :	QSL/WI/M/21
------------------------------	------------------------	-------------

### Standard & Major Equipment (s) Used for calibration:

S.No.	Instrument Name	Make	S. No./Model No.	Certificate No.	Calibrated By	Valid Upto
1.	Std. Weight Box	Weightronics	---	WMCL/F/2023-06/2469	Weightronics	08.06.2025

### Calibration Result:

UUC Reading in µl	Standard Reading in µl	Error in µl	Uncertainty of Measurement:
100	100.221	-0.221	± 0.9 µl
200	200.143	-0.143	± 0.9 µl
500	500.212	-0.212	± 0.9 µl
800	800.328	-0.328	± 0.9 µl
1000	1000.464	-0.464	± 0.9 µl

(The reported expanded Uncertainty in measurement is stated as the standard uncertainty in measurement multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95%)

### Note:

- \*The calibration results reported in this calibration certificate are valid at the time of & under stated conditions of measurement.
- \*This certificate cannot be reproduced except in full without our prior permission in writing.
- \*This certificate refers only to the particular item (s) calibrated.
- \*Laboratory standard are traceable to national standards.
- \*UUC: Unit Under Calibration

Prepared By

Customer Service Cell (Gajendra Singh)

Approved By

CEO (Mridul Kohli)



QSL -2G-314

All business are undertaken subject to company standard general business conditions, which are available on request.