

**PIPETTE CALIBRATION REPORT**

Product code: 4642090  
 Description: Finnpiquette F2 100-1000µl  
 Pipette SN: UW08024  
 Tip: Finntip Flex 1000  
 Nr. of channels: 1

Module SN: -

Manufacturer: Thermo Fisher Scientific, India Pvt. Ltd., D-96, C Road, MIDC Satpur, Nashik - 422007  
 Tested by: 10222434

Date: 12/6/23  
 Balance: Mettler XP205  
 Environment device: Almemo 2590-4AS  
 Air temperature: 26.34 °C  
 Humidity: 46.2 %  
 Z-factor: 1.0042

Time: 4:45:06 PM  
 Balance SN: B115129036, Calibrated 22.11.22  
 Env dev SN: H19070544, Calibrated 08.10.22  
 Water temperature: 25.87 °C  
 Air pressure: 932.5 hPa

Volume:	100 µl	1000 µl
1	99.7	998.1
2	99.9	999.0
3	100.1	999.2
4	100.0	997.8
5	100.0	998.0

Mean	100.0 µl	998.4 µl
A	0.0 µl	-1.6 µl
s	0.2 µl	0.6 µl
Acc%	-0.05	-0.16
CV%	0.17	0.06

**PASSED**

**PASSED**

**Manufacturer specifications:**

	Min.volume:	Max.volume:
	100 µl	1000 µl
A±	1 µl	6 µl
s	0.6 µl	2 µl
Acc%±	1.00	0.60
CV%	0.60	0.20

Manufacturer specifications meet the requirements of ISO 8655