

CALIBRATION CERTIFICATE

Customer Name : National Research Institute of Unani
Medicine for Skin Disorders

Date: **07.06.23**

Equipment Name : SWELAB ALFA

SI.No: 30938

Sl.	DC POWER SUPPLY	TOLERANCE	ACTUAL VALUE	REMARKS
1	+ 24 V	+/- 0.5 V	+24.1 V	
2	+ 20 V	+/- 0.5 V	+20.2 V	
3	+ 5.1 V	+/- 0.2 V	+ 5.2 V	
4	+ 5.0 V	+/- 0.2 V	+ 5.0 V	
5	- 5.0 V	+/- 0.2 V	-5.0 V	
6	+ 3.35 V	+/- 0.2 V	+ 3.34 V	
7	HGB Voltage in Software	3200 - 3800mv	3435mv	

DiaSys Diagnostics India Private Limited

Plot No. A- 821, T.T.C, MIDC, Mahape, Navi Mumbai – 400710. Maharashtra, India.

For feedback/queries contact customer care at: Toll Free number: 1800 120 1447 (India), Email ID : Helpdesk.Service@diasys.de

PERFORMANCE CHECK

Sl	MODULE NAME	CHECK	REMARKS
1	LEVEL DETECTOR CHECK	DONE	-
2	NOISE TEST	DONE	-
3	VALVE / PUMP CHECK	DONE	-
4	TOUCHSCREEN CHECK	DONE	-
5	BLOOD DETECTOR CHECK	DONE	-
6	SHEARVALVE INITIALISATION	DONE	-
7	BACKGROUND CHECK	DONE	-

Actual Run : Performed Calibration and Control for all Parameters and found Ok.

Repeatability Tested & found Ok.

Next Calibration : To be done on 06.06.2024

K. Dagedra
Kumar

DiaSys Diagnostics India Private Limited

Plot No. A- 821, T.T.C, MIDC, Mahape, Navi Mumbai – 400710. Maharashtra, India.

For feedback/queries contact customer care at: Toll Free number: 1800 120 1447 (India), Email ID : Helpdesk.Service@diasys.de