



RML- Quality Assurance Program (RML-QAP)



Lab Code No. **1987**

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 12

ROUND: 6

TOTAL PARTICIPANT: 498

DATE: 29/12/2023

Parameter	Total Responses	Your Result	All Lab Result	Consensus %	Remarks
S1- HIV	398	Non- Reactive	Non- Reactive : 372 Reactive : 26	93.4%	Within Consensus
S2- Dengue IgG	354	Non- Reactive	Non- Reactive : 338 Reactive : 16	95.4%	Within Consensus

Legends:	(*) Excluded From Consensus Group	{NR} Not Reported	(#) Late Result Submission
----------	-----------------------------------	-------------------	----------------------------

Chief Coordinator


Dr. Sanjay Mehrotra

Checked By:

Prepared by: SV

Programme Director

Dr. Bandana Mehrotra



End of Report

Doc. No.: ASS/FR/05/R-01/Dt:05.01.2022

Page 1 of 1

Address: B-171, Nirala Nagar, Lucknow-226020. Ph.: 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555
Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

RML
RESEARCH FOUNDATION

Continuous Efforts And Execution Leads To Quality Excellence



RML- Quality Assurance Program (RML-QAP)



Lab Code No. **1987**

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 12

ROUND: 5

TOTAL PARTICIPANT: 487


DATE: 31/10/2023


Parameter	Total Responses	Your Result	All Lab Result	Consensus %	Remarks
S1- Anti HCV	396	Reactive	-	-	RML Expert Opinion is Reactive (Method ECLIA, ELISA, ICT)
S2- Dengue IgM	362	Non- Reactive	Non- Reactive : 354 Reactive : 08	97.7%	Within Consensus

Legends:	(*) Excluded From Consensus Group	{NR} Not Reported	(#) Late Result Submission
----------	-----------------------------------	-------------------	----------------------------

*Note: Consensus for S1-Anti HCV has not been reached.
Expert Opinion is given.


Chief Coordinator


Dr. Sanjay Mehrotra

Checked By: 

Prepared by: SV

Programme Director


Dr. Bandana Mehrotra



End of Report


RML
Research Foundation

Address: B-171, Nirala Nagar, Lucknow-226020. Ph.: 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Page 1 of 1

Continuous Efforts And Execution Leads To Quality Excellence



RML- Quality Assurance Program (RML-QAP)



Lab Code No. **1987**

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 12

ROUND: 4

TOTAL PARTICIPANT: 475

DATE: 31/08/2023

Parameter	Total Responses	Your Result	All Lab Result	Consensus %	Remarks
S1- Anti HCV	-	-	-	-	-
S2- HBsAg	412	Negative	Positive : 82 Negative : 330	80%	Within Consensus

Legends:	(*) Excluded From Consensus Group	{NR} Not Reported	(#) Late Result Submission
----------	-----------------------------------	-------------------	----------------------------

*Note: Consensus for S1-Anti HCV has not been reached, hence Parameter S1- Anti HCV for Round -4 stands deferred.

Chief Coordinator


Dr. Sanjay Mehrotra

Checked By:

Prepared by: NT

Programme Director

Dr. Bandana Mehrotra



** End of Report **

RML

Doc. No.: ASS/FR/05/R.01/Dt. 05.01.2022

Address: B-171, Nirala Nagar, Lucknow-226020 Ph: 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Page 1 of 1

Continuous Efforts And Execution Leads To Quality Excellence



RML- Quality Assurance Program (RML-QAP)



Lab Code No. 1987

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 12

ROUND: 3

TOTAL PARTICIPANT: 444

DATE: 30/06/2023

Parameter	Total Responses	Your Result	All Lab Result	Consensus %	Remarks
S1- DENGUE IgM	328	NR	Negative : 324 Positive : 4	99%	-
S2- CRP	322	NR	Negative : 94 Positive : 228	71%	-

Legends:	(*) Excluded From Consensus Group	{NR} Not Reported	(#) Late Result Submission
----------	-----------------------------------	-------------------	----------------------------

Chief Coordinator

Dr. Sanjay Mehrotra

Checked By: *A*


Prepared by: NT

Programme Director

Dr. Bandana Mehrotra


kit was changed and method shifted to quantitative turbidometry

End of Report



Doc. No.: ASS / FR / 05 / R 01 / Dt.: 05.01.2022

Page 1 of 1



Address: B-171, Nirala Nagar, Lucknow-226020. Ph.: 4034100, 4077180. Fax: (0522)2788555
Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence



RML – Quality Assurance Program (RML – QAP)



Lab Code No. **1987**

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 12

ROUND: 2

TOTAL PARTICIPANT: 429

DATE: 29/04/2023

Parameter	Total Responses	Your Result	All Lab Result	Consensus %	Remarks
S1- ASO	259	Negative	Negative : 257 Positive : 2	99.2%	Within Consensus
S2- TYPHOID	366	Non- Reactive	Non-Reactive : 359 Reactive : 7	98.0%	Within Consensus

Chief Coordinator


Dr. Sanjay Mehrotra

Checked By:

Prepared by: SV

Programme Director

Dr. Bandana Mehrotra



End of Report

Page 1 of 1

Doc. No. / ASS / FR / 05 / R 01 / Dt.: 05.01.2022

Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph. : 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence

RML- Quality Assurance Program (RML-QAP)



Lab Code No. 1987

BASIC SEROLOGY FINAL RESULT ASSESSMENT

CYCLE NO. : 12

ROUND: 1

TOTAL PARTICIPANT: 387

DATE: 28/02/2023

Parameter	Total Responses	Your Result	All Lab Result	Consensus %	Remarks
S1- RF	276	Negative	Negative : 271 Positive : 5	98%	Within Consensus
S2- RPR	305	Non- Reactive	Non-Reactive : 304 Reactive : 1	99.6%	Within Consensus

Chief Coordinator

Dr. Sanjay Mehrotra

NO CAPA
NPL



Programme Director

Dr. Bandana Mehrotra

Checked By:

Prepared by: SV

End of Report

Doc. No.: ASS / FR / 05 / R 01 / Dt.: 05.01.2022

Page 1 of 1



Address: B-171, Nirala Nagar, Lucknow-226020. Ph.: 4034100-130 (30 Lines), 4077180, 2788444 Fax : (0522)2788555
Email: rmlresearchfoundation@gmail.com Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence