



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029

*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

**EQAP CODE No. :** 2977

**Distribution No.:** 161-G

**Month/Year:** September/2023

**Instrument ID:** Mindray BC6000 Sr.no  
TU-14001804

**Name & Contact No. of PT Co-ordinator:** Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

**Date of issue & status of the report:** 27-12-2023[provisional].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	4.87	4.53	9.4	10.15	0.042	-0.78	0.34	0.1	0.009	2.16
RBC x10 <sup>6</sup> /µl	1	4.12	4.07	8.19	8.03	0.010	0.67	0.05	0.04	0.003	0.27
Hb g/dl	1	12.6	12.6	25.2	25.2	0.030	0.00	0	0.1	0.009	-0.84
HCT%	1	43.4	42.7	86.1	78.4	0.203	1.35	0.7	0.4	0.028	0.81
MCV-fl	1	105.4	105	210.4	194.9	0.388	1.42	0.4	0.2	0.022	0.67
MCH-Pg	1	31.1	30.6	61.7	62.7	0.076	-0.56	0.5	0.3	0.019	0.67
MCHC-g/dl	1	29.5	29	58.5	64.1	0.159	-1.45	0.5	0.3	0.021	0.67
Plt. x10 <sup>3</sup> /µl	1	183	183	366	420	1.689	-1.32	0	6	0.359	-1.16
Retic %	2	7	7	14	14.65	0.279	-0.09	0	0.5	0.034	-0.84

### P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
<b>DLC%</b>	3 Nrbcs=2 , Poly=3 L=93, E=1, Mono/Promono=0 , B1=0 P.M.=0, Mye=0, Meta=0, Other=Prolymphocytes -3, Cleaved cells are seen	Lymp: 85-93, Poly: 3-7, Mono: 1-3, nRBC/Blast/Eosino/Promyelo/Myelo/Meta: 0-5
<b>RBC Morphology</b>	3 Predominantly Normocytic Normochromic admixed Microcytic Hypochromic RBCs	Normocytic, Normochromic; Mild: Microcytic , Hypochromic, Anisopoikilocytosis ,Tear Drop Cells.
<b>Diagnosis</b>	3 CHRONIC LYMPHOCYTIC LEUKEMIA	Chronic Lymphoproliferative Disorder

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 161--G	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	258	257	82.88	81.71	2.33	7.78	14.79	10.51
<b>RBC x10<sup>6</sup>/µl</b>	1	258	258	86.05	92.25	8.53	3.1	5.42	4.65
<b>Hb g/dl</b>	1	258	258	84.88	87.6	7.75	4.26	7.37	8.14
<b>HCT%</b>	1	258	257	95.72	93	3.11	2.33	1.17	4.67
<b>MCV-fl</b>	1	258	257	96.89	89.11	2.72	6.23	0.39	4.66
<b>MCH-Pg</b>	1	258	257	87.16	88.33	7	5.45	5.84	6.22
<b>MCHC-g/dl</b>	1	258	257	92.22	89.11	5.84	7	1.94	3.89
<b>Plt. x10<sup>3</sup>/µl</b>	1	258	257	91.83	92.61	4.67	3.11	3.5	4.28
<b>ReticCount%</b>	2	258	232	92.24	90.95	6.47	6.03	1.29	3.02
<b>PS Assessment</b>	3	258	233	Satisfactory :93.8%, Borderline Sat. :1.55%, Unsatisfactory :4.65%					

**\*Comments:**

**1). Among Lab (EQA) : Results acceptable.**

**2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between " $0$  to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ( $0.3*SDPA$ ). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value ( $0.3*SDPA$ ).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----