



OPTIMA TECHNOCAL SYSTEM

Shyamal Vatika, K-303, Nr. TN Rao Collage, Gangotri Park Road, University Road, RAJKOT-360005 (Guj.)

CALIBRATION CERTIFICATE / PERFORMANCE TEST REPORT

Name of Customer : All India Institute Of Medicoal Sciences (AIIMS)
 Address : AIIMS Hospital, At. Khandheri, Tal. Paddhari, Rajkot-Jamnagar Highway,
 City/Town/Place : Khandheri Village - Dist. Rajkot (Gujarat)



Name of Equipment : PIPETTE Certificate No. : OTS/CAL/PIP/AIIMS-RJ/2024/001

Calibrated On Date	Calibration Due Recommended	Total Page of Report
11_01_2024	11_01_2025	1

I. Calibrated for : Volume (Micro Volume)

II. Item Description and Identification of Equipement / Instrument / Device :

Nomenclature	PIPETTE	Serial No. (If any)	SW22137
Make / Mfg.	Thermo Scientific	ID. No. (If any)	FP-F2 / Thermo Scientific
Model Name / No.	Finnpipette F2	Remark/Type (If any)	50 uL (Fixed Volume) / Micro
Least Count Unit	uL	Dept./Location (If any)	Movable / MicroBiology Lab
Environment Temp.	25 ± 5 *C	Relative Humidity	Relative Humidity: 40 ± 15%

III. Standard / Master Used : *Traceable to National / International Standards - NABL/IAS/ISO/NMI etc.

Nomenclature OR Make/Model Name	Sr. No. / ID. No. Detail (If any)	Traceability Calib. Done By	Traceability Certi. Detail / Number	Due Date of Calib. / Test
Digital Weight Balance Analytical	92730 OR OTS/DWB/01 (WENSAR)	UML - Lab (NABL Certified)	Certi. No. : CC - 219722000003811F	01_08_2024

IV. Result :

Range of Volume Set Micro Volume (uL)	Range of Volume Delivered by DUC (uL)	Measured on Standard	Standard Actual Volume Calculated as per Standard Formula	Error in ±	Result / Status
uL	uL	mg	uL	uL	Each
50	50	0.0501	50.1	-0.1	Pass


NOTE :- Reading final value taken as average of Four readings.

V. Other Details:

- Abbreviations: DUC - Device Under Calibration
 - The Results, Readings and Values are at time on calibration or testing under stated conditions of Test.
 - These Documents are for technical information purpose only. Not for advertisement, litigation of judgement.
- * Accepted Absolute Error ± 3%

Remark: The overall uncertainty is 3% as per manufacturer's specification. Hence confidence level is 97%.
 Based on the above result we declare the instrument / device / equipment is as mentioned below:

Final Status of Calibration / Test Performed	:	PASS / OK
----------------------------------------------	---	-----------

Calibrated/Tested By : 
 --- E.O.C. --- Mr. Dharmendra (Engineer)
 ISO 9001 Certified For, Optima TechnoCal System
 Calibration Laboratory



Calibrators Calibrated Under NABL Laboratory

E-mail : optimatechnocalsystem@gmail.com

Cell : +91 84699 67766, +91 95740 00944