



**Metropolis Healthcare Ltd., EQAS**  
SHREE VIGHNAHARTA SUPERSPECIALITY HOSPITAL (836)

**Outlier And Analyte Summary Report**

Outlier Details For Cycle No 230310 and Sample No 03

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit	Z-Score
* Platelet Count	ERBA H 360	0.07	/µl	*
Legend @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : ERBA H 360

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	1.11	1.12	1.71	0.99
@ Total Leucocytes (WBC) count	/c.mm	2300.00	2450.00	-0.94	0.24
@ Haemoglobin	g/dL	2.90	2.85	0.20	0.59
@ MCV (Mean Corpuscular Volume)	fL	134.20	134.20	-0.08	0.63
@ MCH (Mean Corpuscular Hb)	pg	26.60	25.70	-0.58	-0.16
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	19.80	21.00	-0.47	-0.47
* Platelet Count	1000/ $\mu$ L	0.07	61.00	-6.49 *	-0.77
@ Hematocrit	%	14.90	--	0.58	1.33
@ RDW CV%	%	27.30	--	-0.79	-0.76

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score  $\geq$ 3 - Unacceptable

# : Not Evaluated

⌚ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Haematology - CBC	88.89 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070