



Metropolis Healthcare Ltd., EQAS
SHREE VIGHNAHARTA SUPERSPECIALITY HOSPITAL (836)
Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit
X Glucose (urine)	TransAsia Laura urine analyser	Trace	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : TransAsia Laura urine analyser

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	6.5	6,6.5,7
@ Specific Gravity	-	1.015	1.015,1.020,1.025
X Glucose (urine)	-	Trace	+2,+3,+4
@ Protein (Urine)	-	3+	+1,+2,+3
@ Ketone (urine)	-	Negative	Neg
@ Blood (urine)	-	1+	+1,+2,+3
@ Bilirubin (urine)	-	Negative	Neg
@ Urobilinogen (urine)	-	Normal	Nor
@ Nitrite (urine)	-	Present	Pre

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Urine Pregnancy Test	-	Positive	Pos
Legend @ : Acceptable		Total Parameters	10
@ : Acceptable		Not Evaluated Parameters	0
X : Unacceptable		Evaluated Parameters	10
# : Not Evaluated		Outlier Parameters (X)	1
⊙ : Delayed Result Entry		EQAS Score Clinical Path - Urine Routine Chemistry	90.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroil Road, Kurla (W),
Mumbai - 400070