



Metropolis Healthcare Ltd., EQAS
SHREE VIGHNAHARTA SUPERSPECIALITY HOSPITAL (836)
Outlier And Analyte Summary Report

Outlier Details For Cycle No 230310 and Sample No 01

Report Date : 19/12/2023

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ HIV Ab	S/CO	Non-reactive	Non-reactive
@ Typhidot IgM	-	Negative	Negative
@ Widal test - Typhi O antigen (TO)	-	No Agglutination	1:80,NoA
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	1:80,NoA
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	1:80,NoA
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	1:80,NoA

Instrument : CHEM-5Plus

Analyte	Standard Unit	Result Value	Accepted Value
@ Rheumatoid Factor Qualitative	IU/mL	Normal	Normal
Legend @ : Acceptable		Total Parameters	10
@ : Acceptable		Not Evaluated Parameters	0
X : Unacceptable		Evaluated Parameters	10
# : Not Evaluated		Outlier Parameters (X)	0
⊙ : Delayed Result Entry		EQAS Score Immunoassay Extended	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kiroi Road, Kurla (W),
Mumbai - 400070