




# ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



## CALIBRATION CERTIFICATE

Customer name And Address	M/S: Promila Diagnostic Centre, Shanti Complex, Faizabad Road, Near Lekhraj Metro Station Lucknow, U.P.-226016	Service request No. & date	A/02/06/07/23& 25/12/2023
		ULR no.	CC223923000004182F
		Cert. No.	AT23000004182
		Date of Receipt of DUC	25-12-2023
		Date of calibration	12-01-2024
		Date of issue	17-01-2024
		Suggested due date	11-01-2025

Instrument Details			
Instrument name	Dig. Centrifuge	Location	-----
Make/ModelNo.	Remi R-303	Accuracy	-----
Range/Size	0 to 3500RPM	Visual Inspection	OK
Least Count	10 rpm	Calibration Performed at	Site
I.D.No./Sr.No.	PDC/12/ZJCN-09780		

Detail of reference standards & Major equipment's used			
Equipment Name	Digital Tachometer		
Make	LTLutron		
Model/SRNo.	-----/A.C92792		
CertificateNo.	QCTS/000246/10		
Calibration Validity	20/10/2023		
Calibration by	QCTS		

Environmental Condition	Temperature	20±2°C	Calibration Reference	IS:12508
	Relative Humidity	50±10%	Calibration Procedure	CP-37

### Calibration Results

Serial No.	Standard Value in (rpm)	Measured value in (rpm)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	501.5	502.0	±0.2 %rdg.
02.	1005	1000	
03.	2010	2020	
04.	2990	3000	
05.	3500	3495	

### Remarks:

- ❖ (1) Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2) The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k=2 such that the coverage probability corresponds to approximately 95%. (3) The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By  
(Calibration Engg./TM)  
(MAHIPAL)  
Form No.-QF-47



—End of Report—

Approved By  
(QM/TM)  
(NEERAJ TYAGI)  
PageNo. 1 of 1