



# ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)



## CALIBRATION CERTIFICATE

Customer name And Address	M/S: Promila Diagnostic Centre, Shanti Complex, Faizabad Road, Near Lekhraj Metro Station Lucknow, U.P.-226016	Service request No. & date	A/02/06/07/23& 25/12/2023
		ULR no.	CC223923000004183F
		Cert. No.	AT23000004183
		Date of Receipt of DUC	25-12-2023
		Date of calibration	12-01-2024
		Date of issue	17-01-2024
		Suggested due date	11-01-2025

### Instrument Details

Instrument name	Digital ThermoHygrometer	Location	-----
Make/ModelNo.	HTC-1	Accuracy	-----
Range/Size	-50to70°C/10to99%RH	Visual Inspection	OK
Least Count	0.1°C/1 %RH	Calibration Performed at	Lab
I.D.No./Sr.no.	PDC/15/-----		

### Detail of reference standards & Major equipment's used

Equipment Name	SPRT With Indicator	Dig.Humidity Indicator With Sensor
Make	Tempsans & Fluke	Yudian/ Rotronic
Model/SRNo.	PT-100X1&1586A/05640	11F183060/20274168
Certificate No.	CC288822000001839F	CC325722000000967F
Calibration Validity	27/10/2023	01/11/2023
Calibration by	Sigma Test & Research Centre	KLTECHNOLOGY

Environmental Condition	Temperature	(25±4)°C	Calibration Reference	ITS:90(1990)
	Relative Humidity	(50±20)%	Work Instruction	AT/WI-12TH


### Calibration Results

Serial No.	Standard Average Value in(°C)	Measured Average Value In (°C)	Standard Average Value in(%RH) At @25°C	Measured Average Value in(%RH) At @25°C	UncertaintyAt95%C.L. (coverage factor k=2)
01.	10.022	9.9	30.9	30	
02.	20.037	20.1	49.5	50	±0.3°C
03.	28.985	29.0	69.4	70	
04.	38.972	39.2	79.9	80	±2.0% RH
05.	45.856	46.3	91.4	92	

### Remarks:

- ❖ (1)Standard equipment use for calibration are traceable to national/ international standards.
- ❖ (2)Thereportedexpandeduncertaintyofmeasurementisstatedasthestandarduncertaintyofmeasurementmultipliedbycoveragefactork=2 such that the coverage probability corresponds to approximately 95%. (3)The above results are valid at the time of and under the stated conditions measurement. (4) This certificate is refers only to the particular item submitted for calibration. (5)Next calibration due date given as requested by the customer.

Calibrated By  
(Calibration Engg./TM)  
(D.B SAXENA)  
Form No.-QF-47



—End of Report—

Approved By  
(QM/TM)  
(NEERAJ TYAGI)  
Page No. 1 of 1