

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **YUMIZEN H550** bearing serial number: **212YAXH04063** installed at **KRSNAA DIAGNOSTICS LTD, MALERKOTLA, PUNJAB** has calibrated on 26TH APRIL 2023.

Calibrator : ABX MINOCAL
Lot No. : CX482
Expiry Date: 05/06/2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on **25TH APRIL 2024**.

A circular stamp with the text "HORIBA INDIA PVT. LTD." around the perimeter. Overlaid on the stamp is a handwritten signature in black ink.

Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**