



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT (SUMMARY)

PT SCHEME:HbA1c



PC-1047

Cycle No	C5
Ref.No.	NEUQAP608
Sample ID	NEUQAP HbA1c BLOOD/2023/10/C5/S3

Report Date : 28/12/2023

Sample : OCT 2023

All Methods

	Parameters	Labs (n)	Lab Value	Unit	Assigned value	SDPA	Um	Z-Score	Remarks
1	HbA1c.	9	7.15	%	7.60	0.39		1.15	Satisfactory

Authorised Signatory

Dr. Sujay Prasad
Technical Manager and Program coordinator

* denotes adjusted SDPA as per ISO 13528:2015(E)
SDPA - Standard Deviation for Proficiency Assessment

Technical Advisor

Dr Venkatesh

Page 1 of 3

This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560001
+91-9902277355 · qap@naalm.com · www.naalm.com
CIN: U85300KA2018PTC115147



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT (SUMMARY)

PT SCHEME:HbA1c



PC-1047

Cycle No	C5
Ref.No.	NEUQAP608
Sample ID	NEUQAP HbA1c BLOOD/2023/10/C5/S3

Report Date : 28/12/2023

Sample : OCT 2023

Peer group

	Parameters	Labs (n)	Lab Value	Unit	Assigned value	SDPA	Um	Z-Score	Remarks
2	HbA1c. (Immunoturbidimetry)	3	7.15	%					N/A

Interpretation of z scores is as follows:

| Z score | ≤ 1.0 – Good

| Z score | between 1.0 to 2.0 - Satisfactory

| Z score | between 2.0 to 3.0 - Questionable

| Z score | ≥ 3.0 - Unsatisfactory

Authorised Signatory

Dr. Sujay Prasad
Technical Manager and Program coordinator

* denotes adjusted SDPA as per ISO 13528:2015(E)
SDPA - Standard Deviation for Proficiency Assessment

Technical Advisor

Dr Venkatesh

This report is for use only by the intended participant.

Page 2 of 3



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560001
+91-9902277355 · qap@naalm.com · www.naalm.com
CIN: U85300KA2018PTC115147



Neu-QAP

(an EXTERNAL QUALITY ASSURANCE PROGRAMME)

PARTICIPANT FINAL ASSESSMENT REPORT (NORMAL)

PT SCHEME:HbA1c



PC-1047

Cycle No	C5
Ref.No.	NEUQAP608
Sample ID	NEUQAP HbA1c BLOOD/2023/10/C5/S3


Report Date : 28/12/2023

Sample : OCT 2023

HbA1c.

Parameters	Labs(n)	Lab Value	Unit	Assigned value	SDPA	Um	Z-Score	Remarks
All Methods	9	7.15	%	7.60	0.39		1.15	Satisfactory

ZScore



Peer group (Immunoturbidimetry)	3	7.15	%					N/A
------------------------------------	---	------	---	--	--	--	--	-----

Note: Sample generation, homogeneity and stability testing activities are subcontracted to ISO 15189:2012 accredited laboratory . Lyophilisation of samples is subcontracted to ISO 9001:2015 manufacturing facility .

**End of report **

Authorised Signatory

Dr. Sujay Prasad
Technical Manager and Program coordinator

* denotes adjusted SDPA as per ISO 13528:2015(E)
SDPA - Standard Deviation for Proficiency Assessment

Technical Advisor

Dr Venkatesh

Please refer to NEUQAP Web portal for all subcontracted activ
This report is for use only by the intended participant.



An external quality assurance program from
NEUBERG ANAND ACADEMY OF LABORATORY MEDICINE PVT.LTD.

Anand Tower, 54, Bowring Hospital Road,
Shivajinagar, Bangalore - 560001
+91-9902277355 · qap@naalm.com · www.naalm.com
CIN: U85300KA2018PTC115147