

Calibration Certificate

Certificate No.: YEA(C)/ 231223A/I961/001
 ULR No.: CC222123000082833F

Dated : 30.12.2023

Calibrated for : **ICTC, Kawrthah CHC**
 Dist.: Mamit, Mizoram,
 Pin - 796441.

Calibrated on.	Next Calibration Due On.	Page No.	No. of Pages
25.12.2023	25.12.2024	1	2

Description of Instrument


<i>Name</i>	: Glass Thermometer
<i>Identification</i>	: ICTC/KRT/GT-01
<i>Range</i>	: (-)10 to 110 °C
<i>Least Count</i>	: 1 °C
<i>Make</i>	: Omsons
<i>Type</i>	: Mercury



CC-2221

Service Request No. / Order Reference	: 231223A/1961/001-003
Receipt of Instrument	: 23.12.2023
Location where calibration performed	: At permanent laboratory
Applicable Specification of Instrument to be Calibrated	: Not Specified
Condition of Instrument to be Calibrated	: Satisfactory
Basis of Calibration	: Comparison method with reference to YEA/WI/T-01
Environmental Condition during Calibration	: Temp. : 25±4°C RH : 30 to 75%
Equipment used for Calibration	: Standard Data Acquisition Unit & Standard RTD Sensor
Traceability	: To National Standard vide Certificate No: YEA(C)/ 231220/Y001/001 valid till 19.12.2024 TL/023/52.1.1 valid till 11.01.2024
Measurement Uncertainty	: The figures are stated in the result at coverage factor $k = 2$ corresponding to confidence probability of 95% for normal distribution.
Laboratory Identification	: Sticker is affixed for authentication of calibration.

The above Instrument/Equipment/Machine has been calibrated.
 The results are given in the page 2 to 2 ...of the certificate.

Cont.  ...2





YEA

☎ : 033 2677 1411/1412, 2667 1909

YOUNG ENGG. & CALIBRATION SERVICES PVT. LTD.

Admn. Office & Lab : Kamardanga Road (Opp. Tarama Ashram), Ichapur, Howrah - 711 104

E-mail : calibration@yeagroup.org ★ Visit us at : www.yeagroup.org

CIN : U74140WB2005PTC106389 ★ Toll Free Help Desk : 1800 345 9001

Certificate No.: YEA(C)/ 231223A/I961/001

ULR No.: CC222123000082833F

Calibrated By : K. Rout

Date: 30.12.2023

Page 2 of 2 Pages

Calibration Results

Thermal Calibration

Calibration Results of Glass Thermometer :

Sl. No.	Average DUC Value	Average Standard Value	Measurement Uncertainty
	(°C)	(°C)	(± °C)
1	-10	-10.022	0.62
2	1	0.861	
3	52	51.679	
4	80	79.513	
5	110	109.459	

Temperature Scale : ITS - 90

DUC : Device Under Calibration

REMARKS : The above Glass Thermometer has been calibrated and the readings are tabulated above.

NOTES : This certificate refers to the results tabulated at the time of calibration only for the above mentioned item submitted for calibration and this should not be reproduced, other than in full except with prior written approval from the head of calibration laboratory.

Authorised By

A. Das

Manager Calibration

*** End of the Calibration Certificate ***

