

Calibration done
by engineer.

No. 00004

2023/12/ 7 14:12

WBC	7.9	$\times 10^9/\text{ul}$
LY	2.5	$\times 10^9/\text{ul}$
MO	0.5	$\times 10^9/\text{ul}$
GR	4.9	$\times 10^9/\text{ul}$
LY%	32.2	%
MO%	6.2	%
GR%	61.6	%
RBC	4.61	$\times 10^6/\text{ul}$
Hgb	13.0	g/dl
HCT	39.7	%
MCV	86.1	fl
MCH	28.1	pg
MCHC	32.7	g/dl
RDW	+ 14.7	%
PLT	308	$\times 10^3/\text{ul}$
PCT	+0.295	%
MPV	9.6	fl
PDW	- 10.3	fl

AGD Biomedicals Pvt. Ltd.

Mehta Trade Centre, No.1 Shivaji Colony, Sir M. V. Road, Andheri East, Mumbai 400099. India
 TEL: 91 22 2823 1061 / 2823 1066 FAX : 91 22 28224297 email :ho-service@agdbio.com
 visit us at : www.agdbio.com



Sr. No **92012** **SERVICE REPORT**

Customer Name: RARI Hospital		Instrument Model: PCF-210	
Address/Area: Asowadi - Ahmedabad		Version: S-44K	Serial No: 29168
Name Contact Person: Mr Diptesh Bhuvi		Repair Carried Out:	
Phone/ Mob: 9016167776		Workshop <input type="checkbox"/>	Onsite <input checked="" type="checkbox"/>
Instrument status	Warranty <input type="checkbox"/> Free AMC <input type="checkbox"/> AMC <input checked="" type="checkbox"/> Not in AMC <input type="checkbox"/>	Courtesy <input type="checkbox"/>	Paid Visit <input type="checkbox"/>
Type Of Call	Installation <input type="checkbox"/> Preventative Maintenance <input type="checkbox"/> Break Down <input checked="" type="checkbox"/>	APPLICATION <input type="checkbox"/>	
CALL DETAILS			
CALL RECEIVED		RESPONSE	
DATE: 06/12/23	TIME: 18:00	DATE: 07/12/23	TIME: 12:00
Problem Reported :		CALL COMPLETED	
Observation :		DATE: 07/12/23	TIME: 14:30
pm service & pm kit change. → change pm ket. of PCF-210 → Run Start up & check Background - ok → Run R&D Hematology control - level-2 & it's AU value found in the range. → Instrument working Satisfactorily			
Hemat Reagent	CCR	Voltage Details	L-N 230 E-N 00
Lot No:			L-E 230 CVT / UPS Y[] N[]
<input type="checkbox"/> FOLLOWING PARTS HAVE BEEN REPLACED		Spare Billing Details (if Any)	
<input type="checkbox"/> FOLLOWING PARTS NEED TO BE REPLACED PLEASE APPROVE		Customer Remarks	
SPARES/AMC / CMC / VISIT CHARGE			
Description	QTY	COST	Service : Poor <input type="checkbox"/> Average <input type="checkbox"/> Good <input type="checkbox"/> Excellent <input type="checkbox"/>
			Comments:
FOLLOW UP ACTION (if Require)		WE HEREBY AGREED TO PAY SUM OF Rs: _____	
		Payment Collected: Cheque / DD / Others	
Machine is working satisfactorily. <input checked="" type="checkbox"/>			
Engineer Name:	Bhaktad Dodiya	Customer Signature:	
Engineer Signature:	<i>[Signature]</i>	Date:	
Date: 07/12/23		Stamp	
FOR OFFICE USE ONLY			