



AGD Biomedicals (P) Ltd.
Address- Mehta Trade Center, Sir M.V.Road ,
Andheri East, Mumbai-400099
Tel – +91-22-28231061/66
Email : sales@agdbio.com

DOC.NO.- CC/PCE 210/RARI/008/00
DATE: 27 December 2023

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer ERMA PCE-210 bearing serial number 29168 installed at Regional Ayurved Research Institute Ahmedabad,GJ

Verified with Quality control material CBC-3D R & D systems (Normal) .The reports of Blank Cycle, Control values were all found in acceptable range.
Data given to laboratory.
Instrument is ready for routine work.


Please refer our Service Report No. 92012 on dated 27th Dec,2023 for your reference.

Calibration Date: 27th Dec, 2023
Next Calibration Due on: 26th Dec, 2024.

Enclosed :

- 1) IQC Run Data
- 2) Assay Sheet of Control
- 3) Printouts

Authorized Signatory



For AGD BIOMEDICALS Pvt. Ltd.

AGD Biomedicals (P) Ltd.
Mehta Trade Center,
Sir M. V. Road, Andheri East
Mumbai – 400 099 INDIA

P : +91-22-28231061 / 66
+91-22-28257999
E : sales@agdbio.com
W : www.agdbio.com

ISO : 9001 : 2015
ISO : 13485 : 2016
Page 1 of 2

CONTROL DATA

	WBC	RBC	HgB	MCV	PLT
	[*10 ³ /ul]	[*10 ⁶ /ul]	[g/dl]	fL	[*10 ³ /ul]
LOW					
TARGET					
Range					
Normal					
	WBC	RBC	HgB	MCV	PLT
	7.9	4.61	13.0	86.1	308
TARGET	7.3	4.6	12.6	86.5	265
Range	6.3-8.3	4.3-4.9	11.9-13.3	81.5-91.5	190-340
High					
	WBC	RBC	HgB	MCV	PLT
TARGET					
Range					

This is to certify that the Hematology Analyzer ERMA PCE-210 bearing serial number: 29168 installed at Regional Ayurved Research Institute Ahmedabad,GJ. has been Calibrated using the standardreference material CBC-3D control.

Calibration At site Performed by:

Engineer Name :Prahadbhai Dodiya

Designation /Sign : Service Engineer

AGD Biomedicals Pvt Ltd

